

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP070123\
 Data File : PP058689.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Jul 2023 15:14
 Operator : YP\AJ
 Sample : AR1268ICC750
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 02 07:28:48 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sun Jul 02 07:26:32 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.433	3.666	124.1E6	189.6E6	74.509	71.673
2)	SA Decachlor...	10.235	8.718	147.9E6	263.8E6	72.025	73.232
Target Compounds							
41)	L9 AR-1268-1	8.702	7.603	124.2E6	240.3E6	729.803	732.594
42)	L9 AR-1268-2	8.798	7.667	112.4E6	215.3E6	727.267	734.322
43)	L9 AR-1268-3	9.030	7.876	98248121	190.0E6	726.411	733.091
44)	L9 AR-1268-4	9.461	8.163	36584943	67478264	731.892	723.156
45)	L9 AR-1268-5	9.886	8.459	306.7E6	551.4E6	740.446	742.648



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP070123\
Data File : PP058689.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Jul 2023 15:14
Operator : YP\AJ
Sample : AR1268ICC750
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 02 07:28:48 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070123.M
Quant Title : GC EXTRACTABLES
QLast Update : Sun Jul 02 07:26:32 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.433 min
Delta R.T.: 0.000 min
Response: 124112954
Conc: 74.51 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.666 min
Delta R.T.: 0.000 min
Response: 189568092
Conc: 71.67 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.235 min
Delta R.T.: 0.002 min
Response: 147931585
Conc: 72.02 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.718 min
Delta R.T.: 0.000 min
Response: 263780097
Conc: 73.23 ng/ml



#41 AR-1268-1


R.T.: 8.702 min
Delta R.T.: 0.000 min
Response: 124247066
Conc: 729.80 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 7.603 min
Delta R.T.: 0.000 min
Response: 240303349
Conc: 732.59 ng/ml




#42 AR-1268-2

R.T.: 8.798 min
Delta R.T.: 0.001 min
Response: 112400930
Conc: 727.27 ng/ml



#42 AR-1268-2


R.T.: 7.667 min
Delta R.T.: 0.000 min
Response: 215250781
Conc: 734.32 ng/ml



#43 AR-1268-3


R.T.: 9.030 min
Delta R.T.: 0.000 min
Response: 98248121
Conc: 726.41 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 7.876 min
Delta R.T.: 0.000 min
Response: 190012723
Conc: 733.09 ng/ml




#44 AR-1268-4

R.T.: 9.461 min
Delta R.T.: 0.000 min
Response: 36584943
Conc: 731.89 ng/ml



#44 AR-1268-4


R.T.: 8.163 min
Delta R.T.: 0.000 min
Response: 67478264
Conc: 723.16 ng/ml



#45 AR-1268-5

R.T.: 9.886 min
Delta R.T.: 0.000 min
Response: 306687884
Conc: 740.45 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.459 min
Delta R.T.: 0.000 min
Response: 551391938
Conc: 742.65 ng/ml