

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP070123\
 Data File : PP058684.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Jul 2023 13:52
 Operator : YP\AJ
 Sample : AR1254ICC500
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 02 07:04:07 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sun Jul 02 07:01:48 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.433	3.666	84655099	124.8E6	50.000	50.000
2) SA Decachlor...	10.233	8.718	61419601	98120549	50.000	50.000
Target Compounds						
26) L6 AR-1254-1	6.468	5.571	28259951	61459170	500.000	500.000
27) L6 AR-1254-2	6.688	5.720	42205618	55059905	500.000	500.000
28) L6 AR-1254-3	7.056	6.126	43212575	86864112	500.000	500.000
29) L6 AR-1254-4	7.341	6.356	29751186	49820855	500.000	500.000
30) L6 AR-1254-5	7.762	6.775	34628991	79436230	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP070123\
Data File : PP058684.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Jul 2023 13:52
Operator : YP\AJ
Sample : AR1254ICC500
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 02 07:04:07 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070123.M
Quant Title : GC EXTRACTABLES
QLast Update : Sun Jul 02 07:01:48 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

