

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP070123\
 Data File : PP058672.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Jul 2023 10:37
 Operator : YP\AJ
 Sample : AR1242ICC1000
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 02 06:30:44 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sun Jul 02 06:30:36 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.434	3.667	167.3E6	239.1E6	98.316	98.706
2)	SA Decachlor...	10.235	8.720	115.5E6	173.8E6	96.846	96.034
Target Compounds							
16)	L4 AR-1242-1	5.610	4.765	36813726	54839267	954.626	973.408
17)	L4 AR-1242-2	5.633	4.784	54651329	78979803	970.157	977.418
18)	L4 AR-1242-3	5.695	4.961	34062827	41565805	956.362	974.529
19)	L4 AR-1242-4	5.794	5.046	27358304	43134111	958.683	968.803
20)	L4 AR-1242-5	6.534	5.572	25059542	53112013	911.147	953.839



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP070123\
Data File : PP058672.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Jul 2023 10:37
Operator : YP\AJ
Sample : AR1242ICC1000
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 02 06:30:44 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070123.M
Quant Title : GC EXTRACTABLES
QLast Update : Sun Jul 02 06:30:36 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.434 min
Delta R.T.: 0.000 min
Response: 167347830
Conc: 98.32 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC1000




#1 Tetrachloro-m-xylene

R.T.: 3.667 min
Delta R.T.: 0.000 min
Response: 239074400
Conc: 98.71 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.235 min
Delta R.T.: 0.000 min
Response: 115542522
Conc: 96.85 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.720 min
Delta R.T.: 0.000 min
Response: 173812412
Conc: 96.03 ng/ml



#16 AR-1242-1


R.T.: 5.610 min
Delta R.T.: 0.000 min
Response: 36813726
Conc: 954.63 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC1000




#16 AR-1242-1

R.T.: 4.765 min
Delta R.T.: 0.000 min
Response: 54839267
Conc: 973.41 ng/ml




#17 AR-1242-2

R.T.: 5.633 min
Delta R.T.: 0.000 min
Response: 54651329
Conc: 970.16 ng/ml



#17 AR-1242-2


R.T.: 4.784 min
Delta R.T.: 0.000 min
Response: 78979803
Conc: 977.42 ng/ml



#18 AR-1242-3


R.T.: 5.695 min
Delta R.T.: 0.000 min
Response: 34062827
Conc: 956.36 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC1000




#18 AR-1242-3

R.T.: 4.961 min
Delta R.T.: 0.000 min
Response: 41565805
Conc: 974.53 ng/ml




#19 AR-1242-4

R.T.: 5.794 min
Delta R.T.: 0.000 min
Response: 27358304
Conc: 958.68 ng/ml



#19 AR-1242-4


R.T.: 5.046 min
Delta R.T.: 0.000 min
Response: 43134111
Conc: 968.80 ng/ml



#20 AR-1242-5

R.T.: 6.534 min
Delta R.T.: 0.000 min
Response: 25059542
Conc: 911.15 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC1000



#20 AR-1242-5

R.T.: 5.572 min
Delta R.T.: 0.000 min
Response: 53112013
Conc: 953.84 ng/ml