

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080622\
 Data File : PP050276.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Aug 2022 21:34
 Operator : YP\AJ
 Sample : PB146798BL
 Misc :
 ALS Vial : 70 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 08 00:40:23 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080622.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sun Aug 07 23:52:23 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.345	3.592	35447466	70933156	22.022	22.646
2)	SA Decachlor...	10.088	8.554	41104568	35063654	23.343	22.782
Target Compounds							
9)	L2 AR-1221-2	0.000	3.888	0	953854	N.D.	33.404 #
20)	L4 AR-1242-5	0.000	5.485f	0	562575	N.D.	8.093 #
25)	L5 AR-1248-5	0.000	5.485f	0	562575	N.D.	5.666 #
26)	L6 AR-1254-1	0.000	5.485f	0	562575	N.D.	4.257 #
32)	L7 AR-1260-2	0.000	6.311	0	573981	N.D.	4.273 #
38)	L8 AR-1262-3	0.000	7.481	0	423775	N.D.	5.839 #
41)	L9 AR-1268-1	0.000	7.481	0	423775	N.D.	1.929 #
45)	L9 AR-1268-5	9.779f	0.000	3654155	0	5.479	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP080622\
Data File : PP050276.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Aug 2022 21:34
Operator : YP\AJ
Sample : PB146798BL
Misc :
ALS Vial : 70 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 08 00:40:23 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP080622.M
Quant Title : GC EXTRACTABLES
QLast Update : Sun Aug 07 23:52:23 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.345 min
Delta R.T.: -0.001 min
Response: 35447466
Conc: 22.02 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.592 min
Delta R.T.: -0.001 min
Response: 70933156
Conc: 22.65 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.088 min
Delta R.T.: -0.001 min
Response: 41104568
Conc: 23.34 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.554 min
Delta R.T.: -0.001 min
Response: 35063654
Conc: 22.78 ng/ml



#9 AR-1221-2


R.T.: 0.000 min
Exp R.T.: 4.646 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 3.888 min
Delta R.T.: 0.000 min
Response: 953854
Conc: 33.40 ng/ml




#20 AR-1242-5

R.T.: 0.000 min
Exp R.T.: 6.446 min
Response: 0
Conc: N.D.



#20 AR-1242-5


R.T.: 5.485 min
Delta R.T.: 0.025 min
Response: 562575
Conc: 8.09 ng/ml



#25 AR-1248-5


R.T.: 0.000 min
Exp R.T.: 6.445 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 5.485 min
Delta R.T.: -0.015 min
Response: 562575
Conc: 5.67 ng/ml




#26 AR-1254-1

R.T.: 0.000 min
Exp R.T.: 6.379 min
Response: 0
Conc: N.D.



#26 AR-1254-1


R.T.: 5.485 min
Delta R.T.: 0.024 min
Response: 562575
Conc: 4.26 ng/ml



#32 AR-1260-2


R.T.: 0.000 min
Exp R.T.: 7.386 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#32 AR-1260-2

R.T.: 6.311 min
Delta R.T.: -0.013 min
Response: 573981
Conc: 4.27 ng/ml




#38 AR-1262-3

R.T.: 0.000 min
Exp R.T.: 8.596 min
Response: 0
Conc: N.D.



#38 AR-1262-3


R.T.: 7.481 min
Delta R.T.: 0.012 min
Response: 423775
Conc: 5.84 ng/ml



#41 AR-1268-1


R.T.: 0.000 min
Exp R.T.: 8.597 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 7.481 min
Delta R.T.: 0.011 min
Response: 423775
Conc: 1.93 ng/ml



#45 AR-1268-5

R.T.: 9.779 min
Delta R.T.: 0.022 min
Response: 3654155
Conc: 5.48 ng/ml



#45 AR-1268-5

R.T.: 0.000 min
Exp R.T.: 8.309 min
Response: 0
Conc: N.D.