

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100518\
 Data File : PP027640.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 05 Oct 2018 22:28
 Operator : EI\MA
 Sample : J5204-02
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 06 05:58:30 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092418.M
 Quant Title : 8080.M
 QLast Update : Tue Sep 25 01:51:01 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	12.044	12.159	118.2E6	175.8E6	466.236	406.277
Target Compounds							
1) T	Dalapon	3.272	2.945	118.4E6	171.4E6	198.276	315.165 #
3) T	4-Nitroph...	11.683	11.301	2705040	5674993	6.169	28.221 #
6) T	MCP	12.727	0.000	7444043	0	23.337	N.D. #
7) T	MCPA	0.000	12.981	0 52912423	N.D.	39.413	#
8) T	DICHLORPROP	0.000	13.593	0 12016131	N.D.	28.964	#
10) T	Pentachlo...	0.000	14.680	0 2704404	N.D.	0.410	#
11) T	2,4,5-TP ...	15.237	15.306	6730552	3999511	2.306	1.584 #
12) T	2,4,5-T	15.581	15.895	6191294	29241127	2.477	14.776 #
13) T	2,4-DB	16.346	0.000	8669397	0	45.742	N.D. #
14) T	DINOSEB	17.833	16.949	13970793	811819	7.257	0.511 #
15) T	Picloram	17.597	18.265	58729434	64404065	19.113	25.727 #
16) T	DCPA	0.000	18.140	0 11350106	N.D.	3.556	#



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP100518\
Data File : PP027640.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 05 Oct 2018 22:28
Operator : EI\MA
Sample : J5204-02
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampled :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 06 05:58:30 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP092418.M
Quant Title : 8080.M
QLast Update : Tue Sep 25 01:51:01 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-CLPesticides-1 Signal #2 Phase: ZB-CLPesticides-2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 3.272 min
Delta R.T.: -0.040 min
Response: 118430351
Conc: 198.28 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Dalapon

R.T.: 2.945 min
Delta R.T.: -0.038 min
Response: 171416872
Conc: 315.17 ng/ml




#3 4-Nitrophenol

R.T.: 11.683 min
Delta R.T.: 0.059 min
Response: 2705040
Conc: 6.17 ng/ml



#3 4-Nitrophenol


R.T.: 11.301 min
Delta R.T.: -0.005 min
Response: 5674993
Conc: 28.22 ng/ml



#4 2,4-DCAA


R.T.: 12.044 min
Delta R.T.: -0.002 min
Response: 118244092
Conc: 466.24 ng/ml

Instrument :
ECD_P
ClientSampleId :




#4 2,4-DCAA

R.T.: 12.159 min
Delta R.T.: 0.000 min
Response: 175848017
Conc: 406.28 ng/ml




#6 MCP

R.T.: 12.727 min
Delta R.T.: 0.023 min
Response: 7444043
Conc: 23.34 ug/ml



#6 MCP


R.T.: 0.000 min
Exp R.T. : 12.656 min
Response: 0
Conc: N.D.



#7 MCPA


R.T.: 0.000 min
Exp R.T.: 12.948 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#7 MCPA

R.T.: 12.981 min
Delta R.T.: -0.063 min
Response: 52912423
Conc: 39.41 ug/ml




#8 DICHLORPROP

R.T.: 0.000 min
Exp R.T.: 13.563 min
Response: 0
Conc: N.D.



#8 DICHLORPROP


R.T.: 13.593 min
Delta R.T.: 0.008 min
Response: 12016131
Conc: 28.96 ng/ml



#10 Pentachlorophenol


R.T.: 0.000 min
Exp R.T.: 14.330 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :




#10 Pentachlorophenol

R.T.: 14.680 min
Delta R.T.: 0.000 min
Response: 2704404
Conc: 0.41 ng/ml




#11 2,4,5-TP (SILVEX)

R.T.: 15.237 min
Delta R.T.: 0.044 min
Response: 6730552
Conc: 2.31 ng/ml



#11 2,4,5-TP (SILVEX)


R.T.: 15.306 min
Delta R.T.: 0.055 min
Response: 3999511
Conc: 1.58 ng/ml



#12 2,4,5-T


R.T.: 15.581 min
Delta R.T.: -0.004 min
Response: 6191294
Conc: 2.48 ng/ml

Instrument :
ECD_P
ClientSampleId :




#12 2,4,5-T

R.T.: 15.895 min
Delta R.T.: 0.002 min
Response: 29241127
Conc: 14.78 ng/ml




#13 2,4-DB

R.T.: 16.346 min
Delta R.T.: 0.000 min
Response: 8669397
Conc: 45.74 ng/ml



#13 2,4-DB


R.T.: 0.000 min
Exp R.T. : 16.498 min
Response: 0
Conc: N.D.



#14 DINOSEB


R.T.: 17.833 min
Delta R.T.: 0.000 min
Response: 13970793
Conc: 7.26 ng/ml

Instrument :
ECD_P
ClientSampleId :




#14 DINOSEB

R.T.: 16.949 min
Delta R.T.: 0.064 min
Response: 811819
Conc: 0.51 ng/ml




#15 Picloram

R.T.: 17.597 min
Delta R.T.: 0.000 min
Response: 58729434
Conc: 19.11 ng/ml



#15 Picloram


R.T.: 18.265 min
Delta R.T.: 0.000 min
Response: 64404065
Conc: 25.73 ng/ml



#16 DCPA

R.T.: 0.000 min
Exp R.T.: 18.197 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :



#16 DCPA

R.T.: 18.140 min
Delta R.T.: -0.005 min
Response: 11350106
Conc: 3.56 ng/ml