

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102723\
 Data File : PP061302.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Oct 2023 16:30
 Operator : YP\AJ
 Sample : AR1254ICC250
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 27 16:44:00 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 27 16:43:48 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.558	3.693	62951247	40581123	25.028	26.157
2)	SA Decachlor...	10.529	8.760	49833362	40577839	26.190	26.595
Target Compounds							
26)	L6 AR-1254-1	6.619	5.601	23452853	21135157	266.407	271.357
27)	L6 AR-1254-2	6.841	5.750	35222274	18844095	264.919	273.097
28)	L6 AR-1254-3	7.212	6.157	36425131	30164284	263.754	266.900
29)	L6 AR-1254-4	7.501	6.387	25899407	18443942	260.319	264.838
30)	L6 AR-1254-5	7.924	6.808	28762632	28355447	259.055	264.599



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102723\
Data File : PP061302.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2023 16:30
Operator : YP\AJ
Sample : AR1254ICC250
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 16:44:00 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102723.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 27 16:43:48 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.558 min
Delta R.T.: 0.000 min
Response: 62951247
Conc: 25.03 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC250




#1 Tetrachloro-m-xylene

R.T.: 3.693 min
Delta R.T.: 0.000 min
Response: 40581123
Conc: 26.16 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.529 min
Delta R.T.: 0.000 min
Response: 49833362
Conc: 26.19 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.760 min
Delta R.T.: 0.000 min
Response: 40577839
Conc: 26.59 ng/ml



#26 AR-1254-1


R.T.: 6.619 min
Delta R.T.: 0.000 min
Response: 23452853
Conc: 266.41 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC250




#26 AR-1254-1

R.T.: 5.601 min
Delta R.T.: 0.000 min
Response: 21135157
Conc: 271.36 ng/ml




#27 AR-1254-2

R.T.: 6.841 min
Delta R.T.: 0.000 min
Response: 35222274
Conc: 264.92 ng/ml



#27 AR-1254-2


R.T.: 5.750 min
Delta R.T.: 0.000 min
Response: 18844095
Conc: 273.10 ng/ml



#28 AR-1254-3


R.T.: 7.212 min
Delta R.T.: 0.000 min
Response: 36425131
Conc: 263.75 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC250




#28 AR-1254-3

R.T.: 6.157 min
Delta R.T.: 0.000 min
Response: 30164284
Conc: 266.90 ng/ml




#29 AR-1254-4

R.T.: 7.501 min
Delta R.T.: 0.000 min
Response: 25899407
Conc: 260.32 ng/ml



#29 AR-1254-4


R.T.: 6.387 min
Delta R.T.: 0.000 min
Response: 18443942
Conc: 264.84 ng/ml



#30 AR-1254-5

R.T.: 7.924 min
Delta R.T.: 0.000 min
Response: 28762632
Conc: 259.05 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC250



#30 AR-1254-5

R.T.: 6.808 min
Delta R.T.: 0.000 min
Response: 28355447
Conc: 264.60 ng/ml