

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102723\
 Data File : PP061313.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Oct 2023 19:30
 Operator : YP\AJ
 Sample : AR1254ICV500
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP102723AR1254

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 27 23:25:58 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102723.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 27 18:44:48 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.556	3.691	128.1E6	79442023	51.840	53.157
2) SA Decachlor...	10.522	8.756	98312984	76736076	52.136	49.968
Target Compounds						
26) L6 AR-1254-1	6.615	5.599	48259836	40296842	544.914	511.360
27) L6 AR-1254-2	6.837	5.748	77280038	35770861	579.410	512.167
28) L6 AR-1254-3	7.209	6.155	74144234	58600900	536.100	516.543
29) L6 AR-1254-4	7.498	6.385	51381325	36186692	516.145	520.653
30) L6 AR-1254-5	7.922	6.805	57355148	55617330	521.163	525.635



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP102723\
Data File : PP061313.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Oct 2023 19:30
Operator : YP\AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP102723AR1254

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 27 23:25:58 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP102723.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 27 18:44:48 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.556 min
Delta R.T.: -0.003 min
Response: 128054773
Conc: 51.84 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP102723AR1254




#1 Tetrachloro-m-xylene

R.T.: 3.691 min
Delta R.T.: -0.003 min
Response: 79442023
Conc: 53.16 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.522 min
Delta R.T.: -0.014 min
Response: 98312984
Conc: 52.14 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.756 min
Delta R.T.: -0.009 min
Response: 76736076
Conc: 49.97 ng/ml



#26 AR-1254-1


R.T.: 6.615 min
Delta R.T.: -0.004 min
Response: 48259836
Conc: 544.91 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP102723AR1254




#26 AR-1254-1

R.T.: 5.599 min
Delta R.T.: -0.002 min
Response: 40296842
Conc: 511.36 ng/ml




#27 AR-1254-2

R.T.: 6.837 min
Delta R.T.: -0.003 min
Response: 77280038
Conc: 579.41 ng/ml



#27 AR-1254-2


R.T.: 5.748 min
Delta R.T.: -0.001 min
Response: 35770861
Conc: 512.17 ng/ml



#28 AR-1254-3


R.T.: 7.209 min
Delta R.T.: -0.002 min
Response: 74144234
Conc: 536.10 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP102723AR1254




#28 AR-1254-3

R.T.: 6.155 min
Delta R.T.: -0.002 min
Response: 58600900
Conc: 516.54 ng/ml




#29 AR-1254-4

R.T.: 7.498 min
Delta R.T.: -0.002 min
Response: 51381325
Conc: 516.14 ng/ml



#29 AR-1254-4


R.T.: 6.385 min
Delta R.T.: -0.002 min
Response: 36186692
Conc: 520.65 ng/ml



#30 AR-1254-5

R.T.: 7.922 min
Delta R.T.: -0.003 min
Response: 57355148
Conc: 521.16 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP102723AR1254



#30 AR-1254-5

R.T.: 6.805 min
Delta R.T.: -0.003 min
Response: 55617330
Conc: 525.63 ng/ml