

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP111123\
 Data File : PP061597.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Nov 2023 09:59
 Operator : YP\AJ
 Sample : AR1242ICC1000
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 12 21:07:54 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP111123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sun Nov 12 21:07:46 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.509	3.776	132.2E6	216.5E6	98.995	98.898
2)	SA Decachlor...	10.349	8.935	82865392	169.2E6	95.732	95.909
Target Compounds							
16)	L4 AR-1242-1	5.690	4.898	27627089	58709887	968.814	966.715
17)	L4 AR-1242-2	5.713	4.918	41761588	79325006	973.008	975.213
18)	L4 AR-1242-3	5.776	5.099	27746375	43669954	969.455	966.943
19)	L4 AR-1242-4	5.875	5.185	22102440	44520160	987.052	952.909
20)	L4 AR-1242-5	6.615	5.719	21320624	54580426	991.148	960.563



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP111123\
Data File : PP061597.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Nov 2023 09:59
Operator : YP\AJ
Sample : AR1242ICC1000
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 12 21:07:54 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP111123.M
Quant Title : GC EXTRACTABLES
QLast Update : Sun Nov 12 21:07:46 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.509 min
Delta R.T.: 0.000 min
Response: 132194607
Conc: 99.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC1000




#1 Tetrachloro-m-xylene

R.T.: 3.776 min
Delta R.T.: 0.000 min
Response: 216494361
Conc: 98.90 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.349 min
Delta R.T.: 0.000 min
Response: 82865392
Conc: 95.73 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.935 min
Delta R.T.: 0.000 min
Response: 169242090
Conc: 95.91 ng/ml



#16 AR-1242-1


R.T.: 5.690 min
Delta R.T.: 0.000 min
Response: 27627089
Conc: 968.81 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC1000




#16 AR-1242-1

R.T.: 4.898 min
Delta R.T.: 0.000 min
Response: 58709887
Conc: 966.72 ng/ml




#17 AR-1242-2

R.T.: 5.713 min
Delta R.T.: 0.000 min
Response: 41761588
Conc: 973.01 ng/ml



#17 AR-1242-2


R.T.: 4.918 min
Delta R.T.: 0.000 min
Response: 79325006
Conc: 975.21 ng/ml



#18 AR-1242-3


R.T.: 5.776 min
Delta R.T.: 0.000 min
Response: 27746375
Conc: 969.45 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC1000




#18 AR-1242-3

R.T.: 5.099 min
Delta R.T.: 0.000 min
Response: 43669954
Conc: 966.94 ng/ml




#19 AR-1242-4

R.T.: 5.875 min
Delta R.T.: 0.000 min
Response: 22102440
Conc: 987.05 ng/ml



#19 AR-1242-4


R.T.: 5.185 min
Delta R.T.: 0.000 min
Response: 44520160
Conc: 952.91 ng/ml



#20 AR-1242-5

R.T.: 6.615 min
Delta R.T.: 0.000 min
Response: 21320624
Conc: 991.15 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC1000



#20 AR-1242-5

R.T.: 5.719 min
Delta R.T.: 0.000 min
Response: 54580426
Conc: 960.56 ng/ml