

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ010920\
 Data File : PQ046083.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Jan 2020 21:21
 Operator : AJ\MA
 Sample : AR1232ICC1600
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1232501

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 10 03:00:36 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ010920CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jan 10 02:58:55 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	5.220	4.356	395.3E6	291.8E6	72.451	76.110
2) SA Decachlor...	11.367	9.785	634.9E6	632.7E6	143.828	154.047
Target Compounds						
11) L3 AR-1232-1	5.650	4.813	146.8E6	102.0E6	1318.754	1373.275
12) L3 AR-1232-2	6.242	5.659	81957151	129.1E6	1352.174	1457.253
13) L3 AR-1232-3	6.558	5.856	166.2E6	58863450	1384.978	1432.874
14) L3 AR-1232-4	6.732	5.948	83136100	54675923	1389.147	1388.186
15) L3 AR-1232-5	6.829	6.137	64050505	59829771	1361.368	1400.469

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ010920\
Data File : PQ046083.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Jan 2020 21:21
Operator : AJ\MA
Sample : AR1232ICC1600
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1232501

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 10 03:00:36 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ010920CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jan 10 02:58:55 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

