

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ011323\
 Data File : PQ060222.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Jan 2023 00:27
 Operator : YP\AJ
 Sample : AR1248CCC400
 Misc :
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR12483056

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 14 01:36:50 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ011323CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Jan 14 01:31:33 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.445	2.802	407.0E6	296.8E6	19.704	19.925
2) SA Decachlor...	8.594	7.564	480.9E6	440.6E6	44.290	39.129
Target Compounds						
21) L5 AR-1248-1	4.541	3.805	171.0E6	118.3E6	371.159	371.026
22) L5 AR-1248-2	4.805	4.028	222.7E6	165.6E6	373.505	368.625
23) L5 AR-1248-3	4.995	4.066	269.0E6	165.2E6	374.164	367.998
24) L5 AR-1248-4	5.387	4.224	306.7E6	202.5E6	373.981	370.043
25) L5 AR-1248-5	5.422	4.598	304.3E6	198.4E6	374.592	371.132

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ011323\
Data File : PQ060222.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Jan 2023 00:27
Operator : YP\AJ
Sample : AR1248CCC400
Misc :
ALS Vial : 46 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR12483056

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 14 01:36:50 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ011323CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Jan 14 01:31:33 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

