

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ012221\
 Data File : PQ051826.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Jan 2021 15:55
 Operator : DD\AJ
 Sample : AR1221ICC200
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 23 05:38:42 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ012221CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Jan 23 05:37:49 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml



System Monitoring Compounds							
1)	SA Tetrachlo...	5.231	4.250	258.7E6	75535031	10.405	10.344
2)	SA Decachlor...	11.426	9.598	496.0E6	155.0E6	19.207	20.687
Target Compounds							
8)	L2 AR-1221-1	5.480	4.507	52800583	16966759	208.611	214.294
9)	L2 AR-1221-2	5.580	4.607	36839274	12158546	213.494	226.451
10)	L2 AR-1221-3	5.668	4.696	125.0E6	40861478	210.785	215.952


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ012221\
Data File : PQ051826.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Jan 2021 15:55
Operator : DD\AJ
Sample : AR1221ICC200
Misc :
ALS Vial : 9 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 23 05:38:42 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ012221CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Jan 23 05:37:49 2021
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 5.231 min
Delta R.T.: -0.001 min
Response: 258727268
Conc: 10.40 ng/ml




#1 Tetrachloro-m-xylene

R.T.: 4.250 min
Delta R.T.: -0.002 min
Response: 75535031
Conc: 10.34 ng/ml




#2 Decachlorobiphenyl

R.T.: 11.426 min
Delta R.T.: 0.000 min
Response: 496049797
Conc: 19.21 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.598 min
Delta R.T.: 0.000 min
Response: 154988966
Conc: 20.69 ng/ml




#8 AR-1221-1

R.T.: 5.480 min
Delta R.T.: -0.002 min
Response: 52800583
Conc: 208.61 ng/ml




#8 AR-1221-1

R.T.: 4.507 min
Delta R.T.: -0.002 min
Response: 16966759
Conc: 214.29 ng/ml




#9 AR-1221-2

R.T.: 5.580 min
Delta R.T.: 0.000 min
Response: 36839274
Conc: 213.49 ng/ml




#9 AR-1221-2

R.T.: 4.607 min
Delta R.T.: -0.001 min
Response: 12158546
Conc: 226.45 ng/ml



#10 AR-1221-3

R.T.: 5.668 min
Delta R.T.: 0.000 min
Response: 124998699
Conc: 210.78 ng/ml



#10 AR-1221-3

R.T.: 4.696 min
Delta R.T.: -0.001 min
Response: 40861478
Conc: 215.95 ng/ml