

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ012324\
 Data File : PQ065002.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Jan 2024 00:29
 Operator : YP\AJ
 Sample : P1187-07
 Misc : AR1262 LOD 40 PPB
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 LOD-MDL-WATER-01-QT1-2024

Manual Integrations
 APPROVED

Reviewed By :Ankita Jodhani 01/24/2024
 Supervised By :Sohil Jodhani 01/24/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 24 01:40:56 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ011724.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jan 18 00:09:27 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.472	2.743	75362588	144.2E6	20.230	21.591
2) SA Decachlor...	8.770	7.519	65135845	167.6E6	21.077	23.060
Target Compounds						
36) L8 AR-1262-1	6.722	5.661	4644461	12616127	25.495m	35.368 #
37) L8 AR-1262-2	7.348	5.953	14427348	21630246	29.974	40.928 #
38) L8 AR-1262-3	7.630	6.479	9831250	19446567	32.539	41.895 #
39) L8 AR-1262-4	7.702	6.539	7414266	32985698	31.687	38.636
40) L8 AR-1262-5	8.223	7.033	4478361	15108804	27.923m	36.538m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ012324\
Data File : PQ065002.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Jan 2024 00:29
Operator : YP\AJ
Sample : P1187-07
Misc : AR1262 LOD 40 PPB
ALS Vial : 36 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT1-2024

Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 01/24/2024
Supervised By :Sohil Jodhani 01/24/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 24 01:40:56 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ011724.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jan 18 00:09:27 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

