

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ012419.M
 Title : GC EXTRACTABLES
 Last Update : Thu Jan 24 06:01:26 2019
 Response Via : Initial Calibration

Calibration Files

500 =PQ036662.D 1000 =PQ036660.D 750 =PQ036661.D
 250 =PQ036663.D 50 =PQ036664.D

Compound		500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	4.603	4.556	4.905	4.436	4.514	4.603 E6	3.90	
2) SA	Decachlorobiphenyl	3.621	3.463	3.511	3.625	4.008	3.646 E6	5.88	
3) L1	AR-1016-1	1.346	1.303	1.393	1.374	1.600	1.403 E5	8.19	
4) L1	AR-1016-2	2.049	2.011	2.110	2.047	2.336	2.111 E5	6.21	
5) L1	AR-1016-3	1.196	1.155	1.226	1.215	1.406	1.240 E5	7.81	
6) L1	AR-1016-4	0.975	0.943	0.995	0.974	1.085	0.994 E5	5.43	
7) L1	AR-1016-5	0.970	0.933	0.985	0.979	1.069	0.987 E5	5.10	
8) L2	AR-1221-1	4.553					4.553 E4	0.00	
9) L2	AR-1221-2	3.367					3.367 E4	0.00	
10) L2	AR-1221-3	1.118					1.118 E5	0.00	
11) L3	AR-1232-1	9.310					9.310 E4	0.00	
12) L3	AR-1232-2	4.700					4.700 E4	0.00	
13) L3	AR-1232-3	9.901					9.901 E4	0.00	
14) L3	AR-1232-4	4.653					4.653 E4	0.00	
15) L3	AR-1232-5	3.432					3.432 E4	0.00	
16) L4	AR-1242-1	1.159					1.159 E5	0.00	
17) L4	AR-1242-2	1.773					1.773 E5	0.00	
18) L4	AR-1242-3	1.037					1.037 E5	0.00	
19) L4	AR-1242-4	8.411					8.411 E4	0.00	
20) L4	AR-1242-5	9.074					9.074 E4	0.00	
21) L5	AR-1248-1	8.986					8.986 E4	0.00	
22) L5	AR-1248-2	1.220					1.220 E5	0.00	
23) L5	AR-1248-3	1.373					1.373 E5	0.00	
24) L5	AR-1248-4	1.569					1.569 E5	0.00	
25) L5	AR-1248-5	1.484					1.484 E5	0.00	
26) L6	AR-1254-1	1.773					1.773 E5	0.00	
27) L6	AR-1254-2	2.605					2.605 E5	0.00	
28) L6	AR-1254-3	2.677					2.677 E5	0.00	
29) L6	AR-1254-4	1.984					1.984 E5	0.00	
30) L6	AR-1254-5	2.042					2.042 E5	0.00	
31) L7	AR-1260-1	1.870	1.816	1.838	1.889	2.325	1.948 E5	10.92	
32) L7	AR-1260-2	2.285	2.219	2.229	2.312	2.503	2.309 E5	4.96	
33) L7	AR-1260-3	1.428	1.372	1.377	1.452	1.582	1.442 E5	5.90	
34) L7	AR-1260-4	1.686	1.632	1.627	1.685	1.883	1.703 E5	6.14	
35) L7	AR-1260-5	3.364	3.323	3.288	3.327	3.512	3.363 E5	2.61	
36) L8	AR-1262-1	2.770					2.770 E5	0.00	
37) L8	AR-1262-2	4.904					4.904 E5	0.00	
38) L8	AR-1262-3	2.080					2.080 E5	0.00	
39) L8	AR-1262-4	2.426					2.426 E5	0.00	
40) L8	AR-1262-5	1.627					1.627 E5	0.00	
41) L9	AR-1268-1	6.308					6.308 E5	0.00	
42) L9	AR-1268-2	5.885					5.885 E5	0.00	
43) L9	AR-1268-3	5.061					5.061 E5	0.00	
44) L9	AR-1268-4	1.948					1.948 E5	0.00	
45) L9	AR-1268-5	1.623					1.623 E6	0.00	

Signal #2 Calibration Files

500 =PQ036662.D 1000 =PQ036660.D 750 =PQ036661.D
 250 =PQ036663.D 50 =PQ036664.D

Compound		500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	2.359	2.318	2.512	2.282	2.510	2.396 E6	4.52	
2) SA	Decachlorobiphenyl	1.832	1.725	1.764	1.862	2.224	1.881 E6	10.58	
3) L1	AR-1016-1	7.213	6.961	7.474	7.440	8.574	7.532 E4	8.20	
4) L1	AR-1016-2	1.083	1.057	1.117	1.099	1.209	1.113 E5	5.23	
5) L1	AR-1016-3	5.407	5.196	5.490	5.504	6.601	5.640 E4	9.77	

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ012419.M
 Title : GC EXTRACTABLES
 Last Update : Thu Jan 24 06:01:26 2019
 Response Via : Initial Calibration

Calibration Files

500 =PQ036662.D 1000 =PQ036660.D 750 =PQ036661.D
 250 =PQ036663.D 50 =PQ036664.D

	Compound	500	1000	750	250	50	Avg		%RSD
6)	L1 AR-1016-4	4.386	4.090	4.409	4.561	5.141	4.517	E4	8.59
7)	L1 AR-1016-5	5.617	5.356	5.679	5.820	6.692	5.833	E4	8.73
8)	L2 AR-1221-1	2.528					2.528	E4	0.00
9)	L2 AR-1221-2	1.877					1.877	E4	0.00
10)	L2 AR-1221-3	6.023					6.023	E4	0.00
11)	L3 AR-1232-1	5.031					5.031	E4	0.00
12)	L3 AR-1232-2	5.364					5.364	E4	0.00
13)	L3 AR-1232-3	2.555					2.555	E4	0.00
14)	L3 AR-1232-4	2.368					2.368	E4	0.00
15)	L3 AR-1232-5	2.526					2.526	E4	0.00
16)	L4 AR-1242-1	6.280					6.280	E4	0.00
17)	L4 AR-1242-2	9.333					9.333	E4	0.00
18)	L4 AR-1242-3	4.679					4.679	E4	0.00
19)	L4 AR-1242-4	4.569					4.569	E4	0.00
20)	L4 AR-1242-5	5.598					5.598	E4	0.00
21)	L5 AR-1248-1	4.932					4.932	E4	0.00
22)	L5 AR-1248-2	6.488					6.488	E4	0.00
23)	L5 AR-1248-3	6.627					6.627	E4	0.00
24)	L5 AR-1248-4	7.942					7.942	E4	0.00
25)	L5 AR-1248-5	7.901					7.901	E4	0.00
26)	L6 AR-1254-1	1.339					1.339	E5	0.00
27)	L6 AR-1254-2	1.107					1.107	E5	0.00
28)	L6 AR-1254-3	1.800					1.800	E5	0.00
29)	L6 AR-1254-4	1.320					1.320	E5	0.00
30)	L6 AR-1254-5	1.502					1.502	E5	0.00
31)	L7 AR-1260-1	1.108	1.067	1.081	1.142	1.342	1.148	E5	9.76
32)	L7 AR-1260-2	1.359	1.319	1.316	1.393	1.642	1.406	E5	9.67
33)	L7 AR-1260-3	1.260	1.226	1.228	1.272	1.422	1.282	E5	6.31
34)	L7 AR-1260-4	8.608	8.238	8.262	8.717	9.709	8.707	E4	6.87
35)	L7 AR-1260-5	2.186	2.118	2.102	2.200	2.310	2.183	E5	3.78
36)	L8 AR-1262-1	1.678					1.678	E5	0.00
37)	L8 AR-1262-2	3.028					3.028	E5	0.00
38)	L8 AR-1262-3	1.123					1.123	E5	0.00
39)	L8 AR-1262-4	2.157					2.157	E5	0.00
40)	L8 AR-1262-5	9.405					9.405	E4	0.00
41)	L9 AR-1268-1	3.635					3.635	E5	0.00
42)	L9 AR-1268-2	3.313					3.313	E5	0.00
43)	L9 AR-1268-3	2.726					2.726	E5	0.00
44)	L9 AR-1268-4	1.116					1.116	E5	0.00
45)	L9 AR-1268-5	8.660					8.660	E5	0.00

(#) = Out of Range