

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ012424\
 Data File : PQ065031.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Jan 2024 18:58
 Operator : YP\AJ
 Sample : P1187-08
 Misc : AR1262 LOQ 50 PPB
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 LOQ-WATER-02-QT1-2024

Manual Integrations
 APPROVED

Reviewed By :Ankita Jodhani 01/25/2024
 Supervised By :Sohil Jodhani 01/25/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 25 01:07:32 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ011724.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jan 18 00:09:27 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.472	2.743	81874333	159.6E6	21.978	23.904
2) SA Decachlor...	8.778	7.521	70367026	181.7E6	22.769	25.005
Target Compounds						
36) L8 AR-1262-1	6.725	5.661	6694413	16486736	36.748m	46.219 #
37) L8 AR-1262-2	7.354	5.952	19752677	28167364	41.038	53.297 #
38) L8 AR-1262-3	7.635	6.478	12917481	31715299	42.753	68.327 #
39) L8 AR-1262-4	7.709	6.538	9556160	44943928	40.841	52.642 #
40) L8 AR-1262-5	8.231	7.033	6338996	18425651	39.524m	44.559m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ012424\
Data File : PQ065031.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Jan 2024 18:58
Operator : YP\AJ
Sample : P1187-08
Misc : AR1262 LOQ 50 PPB
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
LOQ-WATER-02-QT1-2024

Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 01/25/2024
Supervised By :Sohil Jodhani 01/25/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 25 01:07:32 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ011724.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jan 18 00:09:27 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

