

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ012522\
 Data File : PQ055941.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Jan 2022 11:35
 Operator : AJ\MA
 Sample : N1222-05 10X
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 CLI-2022-3A

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 25 22:31:40 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ011922.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jan 20 13:18:56 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	5.235	4.293	18292673	15780788	1.351	1.404
2)	SA Decachlor...	11.418	9.658	22186956	11981440	1.520	1.502
Target Compounds							
3)	L1 AR-1016-1	6.565	5.576	2829891	4229270	8.092	13.772 #
4)	L1 AR-1016-2	6.587	5.598	357774	844925	0.526	1.356 #
5)	L1 AR-1016-3	6.645	5.789	2246753	2034179	4.942	7.453 #
6)	L1 AR-1016-4	6.786f	5.819	3478732	1265173	9.394	5.014 #
7)	L1 AR-1016-5	7.062	6.053	996781	50647	3.404	0.173 #
8)	L2 AR-1221-1	5.482	4.582f	215457	810735	1.510	7.159 #
9)	L2 AR-1221-2	5.592	4.662	139272	556622	1.270	6.360 #
10)	L2 AR-1221-3	5.682	4.746	184061	281059	0.567	0.996 #
11)	L3 AR-1232-1	5.682	4.746	184061	281059	0.686	1.223 #
12)	L3 AR-1232-2	6.268	5.598	661829	844925	4.464	3.475
13)	L3 AR-1232-3	6.587	5.789	357774	2034179	1.288	19.584 #
14)	L3 AR-1232-4	6.786f	0.000	3478732	0	23.522	N.D. #
15)	L3 AR-1232-5	6.860	6.053	690709	50647	7.305	0.462 #
16)	L4 AR-1242-1	6.565	5.576	2829891	4229270	10.488	18.272 #
17)	L4 AR-1242-2	6.592	5.598	783097	844925	1.496	1.794
18)	L4 AR-1242-3	6.645	5.789	2246753	2034179	6.213	9.795 #
19)	L4 AR-1242-4	6.786f	0.000	3478732	0	11.924	N.D. #
20)	L4 AR-1242-5	7.555	6.435	324926	1058687	1.226	3.921 #
21)	L5 AR-1248-1	6.565	5.576	2829891	4229270	13.514	24.466 #
22)	L5 AR-1248-2	6.860	5.819	690709	1265173	2.157	3.966 #
23)	L5 AR-1248-3	7.062	0.000	996781	0	2.635	N.D. #
24)	L5 AR-1248-4	7.512	6.053	61932	50647	0.146	0.134
25)	L5 AR-1248-5	7.555	6.477	324926	720355	0.706	1.959 #
26)	L6 AR-1254-1	7.471	6.435	651045	1058687	1.598	1.708
27)	L6 AR-1254-2	7.703	6.589	1100277	591471	1.698	1.092 #
28)	L6 AR-1254-3	8.081	7.015	2475033	897189	3.165	1.040 #
29)	L6 AR-1254-4	8.385	7.256	693473	518128	1.224	0.829 #
30)	L6 AR-1254-5	8.807	7.683	2524761	882203	4.000	1.163 #
31)	L7 AR-1260-1	8.247	7.151	2907446	602413	6.049	1.100 #
32)	L7 AR-1260-2	8.527	7.345	3952058	844520	6.709	1.272 #
33)	L7 AR-1260-3	8.876	7.487f	1283452	2233542	2.567	3.629 #
34)	L7 AR-1260-4	9.122	7.985	2904251	395183	4.986	0.782 #
35)	L7 AR-1260-5	9.479	8.237	2294843	727009	1.746	0.612 #
36)	L8 AR-1262-1	8.876	7.683	1283452	882203	1.792	2.427 #
37)	L8 AR-1262-2	9.479	8.237	2294843	727009	1.534	0.550 #
38)	L8 AR-1262-3	9.796	8.523	329051	142758	0.428	0.288 #
39)	L8 AR-1262-4	9.912	8.578	382880	449260	0.579	0.487
40)	L8 AR-1262-5	10.599f	9.104	2017391	275442	2.949	0.691 #
41)	L9 AR-1268-1	9.796	8.523	329051	142758	0.174	0.097 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ012522\
 Data File : PQ055941.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Jan 2022 11:35
 Operator : AJ\MA
 Sample : N1222-05 10X
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 CLI-2022-3A

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 25 22:31:40 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ011922.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jan 20 13:18:56 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	9.912	8.578	382880	449260	0.192	0.339 #
43) L9	AR-1268-3	10.135	8.795	100619	180280	0.059	0.163 #
44) L9	AR-1268-4	10.599f	9.104	2017391	275442	2.706	0.627 #
45) L9	AR-1268-5	11.052	9.386	635200	144117	0.116	0.044 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ012522\
Data File : PQ055941.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Jan 2022 11:35
Operator : AJ\MA
Sample : N1222-05 10X
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
CLI-2022-3A

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 25 22:31:40 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ011922.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jan 20 13:18:56 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 5.235 min
Delta R.T.: 0.000 min
Response: 18292673
Conc: 1.35 ng/ml

Instrument :
ECD_Q
ClientSampleId :
CLI-2022-3A




#1 Tetrachloro-m-xylene

R.T.: 4.293 min
Delta R.T.: -0.001 min
Response: 15780788
Conc: 1.40 ng/ml




#2 Decachlorobiphenyl

R.T.: 11.418 min
Delta R.T.: -0.004 min
Response: 22186956
Conc: 1.52 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.658 min
Delta R.T.: -0.008 min
Response: 11981440
Conc: 1.50 ng/ml



#3 AR-1016-1


R.T.: 6.565 min
Delta R.T.: 0.000 min
Response: 2829891
Conc: 8.09 ng/ml

Instrument :
ECD_Q
ClientSampleId :
CLI-2022-3A




#3 AR-1016-1

R.T.: 5.576 min
Delta R.T.: 0.008 min
Response: 4229270
Conc: 13.77 ng/ml




#4 AR-1016-2

R.T.: 6.587 min
Delta R.T.: -0.001 min
Response: 357774
Conc: 0.53 ng/ml



#4 AR-1016-2


R.T.: 5.598 min
Delta R.T.: 0.009 min
Response: 844925
Conc: 1.36 ng/ml



#5 AR-1016-3


R.T.: 6.645 min
Delta R.T.: -0.009 min
Response: 2246753
Conc: 4.94 ng/ml

Instrument :
ECD_Q
ClientSampleId :
CLI-2022-3A




#5 AR-1016-3

R.T.: 5.789 min
Delta R.T.: 0.008 min
Response: 2034179
Conc: 7.45 ng/ml




#6 AR-1016-4

R.T.: 6.786 min
Delta R.T.: 0.024 min
Response: 3478732
Conc: 9.39 ng/ml



#6 AR-1016-4


R.T.: 5.819 min
Delta R.T.: -0.008 min
Response: 1265173
Conc: 5.01 ng/ml



#7 AR-1016-5


R.T.: 7.062 min
Delta R.T.: -0.012 min
Response: 996781
Conc: 3.40 ng/ml

Instrument :
ECD_Q
ClientSampleId :
CLI-2022-3A




#7 AR-1016-5

R.T.: 6.053 min
Delta R.T.: -0.006 min
Response: 50647
Conc: 0.17 ng/ml




#8 AR-1221-1

R.T.: 5.482 min
Delta R.T.: -0.007 min
Response: 215457
Conc: 1.51 ng/ml



#8 AR-1221-1


R.T.: 4.582 min
Delta R.T.: 0.023 min
Response: 810735
Conc: 7.16 ng/ml



#9 AR-1221-2


R.T.: 5.592 min
Delta R.T.: 0.004 min
Response: 139272
Conc: 1.27 ng/ml

Instrument :
ECD_Q
ClientSampleId :
CLI-2022-3A




#9 AR-1221-2

R.T.: 4.662 min
Delta R.T.: 0.003 min
Response: 556622
Conc: 6.36 ng/ml




#10 AR-1221-3

R.T.: 5.682 min
Delta R.T.: 0.006 min
Response: 184061
Conc: 0.57 ng/ml



#10 AR-1221-3


R.T.: 4.746 min
Delta R.T.: -0.004 min
Response: 281059
Conc: 1.00 ng/ml



#11 AR-1232-1


R.T.: 5.682 min
Delta R.T.: 0.006 min
Response: 184061
Conc: 0.69 ng/ml

Instrument :
ECD_Q
ClientSampleId :
CLI-2022-3A




#11 AR-1232-1

R.T.: 4.746 min
Delta R.T.: -0.003 min
Response: 281059
Conc: 1.22 ng/ml




#12 AR-1232-2

R.T.: 6.268 min
Delta R.T.: 0.002 min
Response: 661829
Conc: 4.46 ng/ml



#12 AR-1232-2


R.T.: 5.598 min
Delta R.T.: 0.010 min
Response: 844925
Conc: 3.47 ng/ml



#13 AR-1232-3


R.T.: 6.587 min
Delta R.T.: -0.001 min
Response: 357774
Conc: 1.29 ng/ml

Instrument :
ECD_Q
ClientSampleId :
CLI-2022-3A




#13 AR-1232-3

R.T.: 5.789 min
Delta R.T.: 0.008 min
Response: 2034179
Conc: 19.58 ng/ml




#14 AR-1232-4

R.T.: 6.786 min
Delta R.T.: 0.023 min
Response: 3478732
Conc: 23.52 ng/ml



#14 AR-1232-4


R.T.: 0.000 min
Exp R.T.: 5.872 min
Response: 0
Conc: N.D.



#15 AR-1232-5


R.T.: 6.860 min
Delta R.T.: 0.006 min
Response: 690709
Conc: 7.30 ng/ml

Instrument :
ECD_Q
ClientSampleId :
CLI-2022-3A




#15 AR-1232-5

R.T.: 6.053 min
Delta R.T.: -0.005 min
Response: 50647
Conc: 0.46 ng/ml




#16 AR-1242-1

R.T.: 6.565 min
Delta R.T.: 0.000 min
Response: 2829891
Conc: 10.49 ng/ml



#16 AR-1242-1


R.T.: 5.576 min
Delta R.T.: 0.008 min
Response: 4229270
Conc: 18.27 ng/ml



#17 AR-1242-2


R.T.: 6.592 min
Delta R.T.: 0.002 min
Response: 783097
Conc: 1.50 ng/ml

Instrument :
ECD_Q
ClientSampleId :
CLI-2022-3A




#17 AR-1242-2

R.T.: 5.598 min
Delta R.T.: 0.009 min
Response: 844925
Conc: 1.79 ng/ml




#18 AR-1242-3

R.T.: 6.645 min
Delta R.T.: -0.010 min
Response: 2246753
Conc: 6.21 ng/ml



#18 AR-1242-3


R.T.: 5.789 min
Delta R.T.: 0.007 min
Response: 2034179
Conc: 9.80 ng/ml



#19 AR-1242-4


R.T.: 6.786 min
Delta R.T.: 0.023 min
Response: 3478732
Conc: 11.92 ng/ml

Instrument :
ECD_Q
ClientSampleId :
CLI-2022-3A




#19 AR-1242-4

R.T.: 0.000 min
Exp R.T. : 5.873 min
Response: 0
Conc: N.D.




#20 AR-1242-5

R.T.: 7.555 min
Delta R.T.: 0.008 min
Response: 324926
Conc: 1.23 ng/ml



#20 AR-1242-5


R.T.: 6.435 min
Delta R.T.: -0.006 min
Response: 1058687
Conc: 3.92 ng/ml



#21 AR-1248-1


R.T.: 6.565 min
Delta R.T.: 0.000 min
Response: 2829891
Conc: 13.51 ng/ml

Instrument :
ECD_Q
ClientSampleId :
CLI-2022-3A




#21 AR-1248-1

R.T.: 5.576 min
Delta R.T.: 0.008 min
Response: 4229270
Conc: 24.47 ng/ml




#22 AR-1248-2

R.T.: 6.860 min
Delta R.T.: 0.006 min
Response: 690709
Conc: 2.16 ng/ml



#22 AR-1248-2


R.T.: 5.819 min
Delta R.T.: -0.007 min
Response: 1265173
Conc: 3.97 ng/ml



#23 AR-1248-3


R.T.: 7.062 min
Delta R.T.: -0.011 min
Response: 996781
Conc: 2.63 ng/ml

Instrument :
ECD_Q
ClientSampleId :
CLI-2022-3A




#23 AR-1248-3

R.T.: 0.000 min
Exp R.T. : 5.872 min
Response: 0
Conc: N.D.




#24 AR-1248-4

R.T.: 7.512 min
Delta R.T.: 0.008 min
Response: 61932
Conc: 0.15 ng/ml



#24 AR-1248-4


R.T.: 6.053 min
Delta R.T.: -0.005 min
Response: 50647
Conc: 0.13 ng/ml



#25 AR-1248-5


R.T.: 7.555 min
Delta R.T.: 0.009 min
Response: 324926
Conc: 0.71 ng/ml

Instrument :
ECD_Q
ClientSampleId :
CLI-2022-3A




#25 AR-1248-5

R.T.: 6.477 min
Delta R.T.: -0.009 min
Response: 720355
Conc: 1.96 ng/ml




#26 AR-1254-1

R.T.: 7.471 min
Delta R.T.: -0.001 min
Response: 651045
Conc: 1.60 ng/ml



#26 AR-1254-1


R.T.: 6.435 min
Delta R.T.: -0.003 min
Response: 1058687
Conc: 1.71 ng/ml



#27 AR-1254-2


R.T.: 7.703 min
Delta R.T.: 0.003 min
Response: 1100277
Conc: 1.70 ng/ml

Instrument :
ECD_Q
ClientSampleId :
CLI-2022-3A




#27 AR-1254-2

R.T.: 6.589 min
Delta R.T.: -0.006 min
Response: 591471
Conc: 1.09 ng/ml




#28 AR-1254-3

R.T.: 8.081 min
Delta R.T.: -0.004 min
Response: 2475033
Conc: 3.16 ng/ml



#28 AR-1254-3


R.T.: 7.015 min
Delta R.T.: -0.006 min
Response: 897189
Conc: 1.04 ng/ml



#29 AR-1254-4


R.T.: 8.385 min
Delta R.T.: 0.003 min
Response: 693473
Conc: 1.22 ng/ml

Instrument :
ECD_Q
ClientSampleId :
CLI-2022-3A




#29 AR-1254-4

R.T.: 7.256 min
Delta R.T.: -0.006 min
Response: 518128
Conc: 0.83 ng/ml




#30 AR-1254-5

R.T.: 8.807 min
Delta R.T.: -0.003 min
Response: 2524761
Conc: 4.00 ng/ml



#30 AR-1254-5


R.T.: 7.683 min
Delta R.T.: -0.005 min
Response: 882203
Conc: 1.16 ng/ml



#31 AR-1260-1


R.T.: 8.247 min
Delta R.T.: -0.004 min
Response: 2907446
Conc: 6.05 ng/ml

Instrument :
ECD_Q
ClientSampleId :
CLI-2022-3A




#31 AR-1260-1

R.T.: 7.151 min
Delta R.T.: -0.004 min
Response: 602413
Conc: 1.10 ng/ml




#32 AR-1260-2

R.T.: 8.527 min
Delta R.T.: 0.009 min
Response: 3952058
Conc: 6.71 ng/ml



#32 AR-1260-2


R.T.: 7.345 min
Delta R.T.: -0.007 min
Response: 844520
Conc: 1.27 ng/ml



#33 AR-1260-3


R.T.: 8.876 min
Delta R.T.: -0.009 min
Response: 1283452
Conc: 2.57 ng/ml

Instrument :
ECD_Q
ClientSampleId :
CLI-2022-3A




#33 AR-1260-3

R.T.: 7.487 min
Delta R.T.: -0.022 min
Response: 2233542
Conc: 3.63 ng/ml




#34 AR-1260-4

R.T.: 9.122 min
Delta R.T.: -0.004 min
Response: 2904251
Conc: 4.99 ng/ml



#34 AR-1260-4


R.T.: 7.985 min
Delta R.T.: -0.007 min
Response: 395183
Conc: 0.78 ng/ml



#35 AR-1260-5


R.T.: 9.479 min
Delta R.T.: 0.004 min
Response: 2294843
Conc: 1.75 ng/ml

Instrument :
ECD_Q
ClientSampleId :
CLI-2022-3A




#35 AR-1260-5

R.T.: 8.237 min
Delta R.T.: -0.003 min
Response: 727009
Conc: 0.61 ng/ml




#36 AR-1262-1

R.T.: 8.876 min
Delta R.T.: -0.006 min
Response: 1283452
Conc: 1.79 ng/ml



#36 AR-1262-1


R.T.: 7.683 min
Delta R.T.: -0.004 min
Response: 882203
Conc: 2.43 ng/ml



#37 AR-1262-2


R.T.: 9.479 min
Delta R.T.: 0.007 min
Response: 2294843
Conc: 1.53 ng/ml

Instrument :
ECD_Q
ClientSampleId :
CLI-2022-3A




#37 AR-1262-2

R.T.: 8.237 min
Delta R.T.: -0.001 min
Response: 727009
Conc: 0.55 ng/ml




#38 AR-1262-3

R.T.: 9.796 min
Delta R.T.: -0.007 min
Response: 329051
Conc: 0.43 ng/ml



#38 AR-1262-3


R.T.: 8.523 min
Delta R.T.: -0.001 min
Response: 142758
Conc: 0.29 ng/ml



#39 AR-1262-4


R.T.: 9.912 min
Delta R.T.: 0.007 min
Response: 382880
Conc: 0.58 ng/ml

Instrument :
ECD_Q
ClientSampleId :
CLI-2022-3A




#39 AR-1262-4

R.T.: 8.578 min
Delta R.T.: -0.012 min
Response: 449260
Conc: 0.49 ng/ml




#40 AR-1262-5

R.T.: 10.599 min
Delta R.T.: -0.016 min
Response: 2017391
Conc: 2.95 ng/ml



#40 AR-1262-5


R.T.: 9.104 min
Delta R.T.: 0.008 min
Response: 275442
Conc: 0.69 ng/ml



#41 AR-1268-1


R.T.: 9.796 min
Delta R.T.: -0.007 min
Response: 329051
Conc: 0.17 ng/ml

Instrument :
ECD_Q
ClientSampleId :
CLI-2022-3A




#41 AR-1268-1

R.T.: 8.523 min
Delta R.T.: -0.001 min
Response: 142758
Conc: 0.10 ng/ml




#42 AR-1268-2

R.T.: 9.912 min
Delta R.T.: 0.006 min
Response: 382880
Conc: 0.19 ng/ml



#42 AR-1268-2


R.T.: 8.578 min
Delta R.T.: -0.011 min
Response: 449260
Conc: 0.34 ng/ml



#43 AR-1268-3


R.T.: 10.135 min
Delta R.T.: -0.012 min
Response: 100619
Conc: 0.06 ng/ml

Instrument :
ECD_Q
ClientSampleId :
CLI-2022-3A




#43 AR-1268-3

R.T.: 8.795 min
Delta R.T.: -0.003 min
Response: 180280
Conc: 0.16 ng/ml




#44 AR-1268-4

R.T.: 10.599 min
Delta R.T.: -0.017 min
Response: 2017391
Conc: 2.71 ng/ml



#44 AR-1268-4


R.T.: 9.104 min
Delta R.T.: 0.009 min
Response: 275442
Conc: 0.63 ng/ml



#45 AR-1268-5

R.T.: 11.052 min
Delta R.T.: -0.010 min
Response: 635200
Conc: 0.12 ng/ml

Instrument :
ECD_Q
ClientSampleId :
CLI-2022-3A



#45 AR-1268-5

R.T.: 9.386 min
Delta R.T.: -0.012 min
Response: 144117
Conc: 0.04 ng/ml