

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ012522\
 Data File : PQ055944.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Jan 2022 12:24
 Operator : AJ\MA
 Sample : N1222-09 10X
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 CLI-2022-5A

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 25 22:32:39 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ011922.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jan 20 13:18:56 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	5.236	4.294	22308258	18662468	1.647	1.660
2)	SA Decachlor...	11.417	9.657	24882856	13186155	1.704	1.653
Target Compounds							
3)	L1 AR-1016-1	6.563	5.581	5269495	3509341	15.069	11.428
4)	L1 AR-1016-2	6.563f	5.581	5269495	3509341	7.751	5.634 #
5)	L1 AR-1016-3	6.642	5.769	9322346	1669026	20.504	6.115 #
6)	L1 AR-1016-4	6.778f	5.815	1672768	1272799	4.517	5.044
7)	L1 AR-1016-5	7.068	6.048	742700	168061	2.536	0.574 #
8)	L2 AR-1221-1	5.498	4.582f	614950	77647	4.309	0.686 #
9)	L2 AR-1221-2	5.598	4.652	379932	213332	3.464	2.437 #
10)	L2 AR-1221-3	5.669	4.748	617925	46938	1.904	0.166 #
11)	L3 AR-1232-1	5.669	4.748	617925	46938	2.304	0.204 #
12)	L3 AR-1232-2	6.281f	5.581	8383631	3509341	56.548	14.432 #
13)	L3 AR-1232-3	6.563f	5.769	5269495	1669026	18.964	16.068
14)	L3 AR-1232-4	6.778f	5.900f	1672768	927634	11.311	8.936
15)	L3 AR-1232-5	6.870f	6.048	1073757	168061	11.356	1.532 #
16)	L4 AR-1242-1	6.563	5.581	5269495	3509341	19.530	15.162
17)	L4 AR-1242-2	6.563f	5.581	5269495	3509341	10.068	7.453 #
18)	L4 AR-1242-3	6.642	5.769	9322346	1669026	25.778	8.037 #
19)	L4 AR-1242-4	6.778f	5.900f	1672768	927634	5.734	4.158 #
20)	L4 AR-1242-5	7.543	6.434	750709	674098	2.832	2.496
21)	L5 AR-1248-1	6.563	5.581	5269495	3509341	25.165	20.302
22)	L5 AR-1248-2	6.870f	5.815	1073757	1272799	3.354	3.990
23)	L5 AR-1248-3	7.068	5.900f	742700	927634	1.963	2.834 #
24)	L5 AR-1248-4	7.502	6.048	241516	168061	0.571	0.446
25)	L5 AR-1248-5	7.543	6.476	750709	227581	1.630	0.619 #
26)	L6 AR-1254-1	7.469	6.434	248742	674098	0.611	1.087 #
27)	L6 AR-1254-2	7.702	6.592	1206676	370446	1.862	0.684 #
28)	L6 AR-1254-3	8.083	7.020	1221031	345629	1.561	0.401 #
29)	L6 AR-1254-4	8.374	7.257	1109263	250185	1.958	0.400 #
30)	L6 AR-1254-5	8.810	7.682	3599207	1030031	5.702	1.358 #
31)	L7 AR-1260-1	8.246	7.150	2648814	627892	5.511	1.147 #
32)	L7 AR-1260-2	8.513	7.345	3860444	870658	6.553	1.312 #
33)	L7 AR-1260-3	8.881	7.530f	973049	333997	1.946	0.543 #
34)	L7 AR-1260-4	9.127	7.986	1763460	633509	3.027	1.254 #
35)	L7 AR-1260-5	9.472	8.232	8666271	1490571	6.594	1.255 #
36)	L8 AR-1262-1	8.881	7.682	973049	1030031	1.358	2.834 #
37)	L8 AR-1262-2	9.472	8.232	8666271	1490571	5.791	1.127 #
38)	L8 AR-1262-3	9.799	8.520	434005	853260	0.564	1.724 #
39)	L8 AR-1262-4	9.914	8.582	756715	1108160	1.144	1.201
40)	L8 AR-1262-5	10.611	9.095	2410789	378376	3.524	0.950 #
41)	L9 AR-1268-1	9.799	8.520	434005	853260	0.230	0.582 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ012522\
Data File : PQ055944.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Jan 2022 12:24
Operator : AJ\MA
Sample : N1222-09 10X
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
CLI-2022-5A

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 25 22:32:39 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ011922.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jan 20 13:18:56 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42)	L9 AR-1268-2	9.914	8.582	756715	1108160	0.379	0.836 #
43)	L9 AR-1268-3	10.166f	8.789	1352145	590959	0.790	0.535 #
44)	L9 AR-1268-4	10.611	9.095	2410789	378376	3.234	0.861 #
45)	L9 AR-1268-5	11.056	9.398	612338	216810	0.111	0.066 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ012522\
Data File : PQ055944.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Jan 2022 12:24
Operator : AJ\MA
Sample : N1222-09 10X
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
CLI-2022-5A

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 25 22:32:39 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ011922.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jan 20 13:18:56 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 5.236 min
Delta R.T.: 0.000 min
Response: 22308258
Conc: 1.65 ng/ml

Instrument :
ECD_Q
ClientSampleId :
CLI-2022-5A




#1 Tetrachloro-m-xylene

R.T.: 4.294 min
Delta R.T.: 0.000 min
Response: 18662468
Conc: 1.66 ng/ml




#2 Decachlorobiphenyl

R.T.: 11.417 min
Delta R.T.: -0.005 min
Response: 24882856
Conc: 1.70 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.657 min
Delta R.T.: -0.009 min
Response: 13186155
Conc: 1.65 ng/ml



#3 AR-1016-1


R.T.: 6.563 min
Delta R.T.: -0.001 min
Response: 5269495
Conc: 15.07 ng/ml

Instrument :
ECD_Q
ClientSampleId :
CLI-2022-5A




#3 AR-1016-1

R.T.: 5.581 min
Delta R.T.: 0.014 min
Response: 3509341
Conc: 11.43 ng/ml




#4 AR-1016-2

R.T.: 6.563 min
Delta R.T.: -0.025 min
Response: 5269495
Conc: 7.75 ng/ml



#4 AR-1016-2


R.T.: 5.581 min
Delta R.T.: -0.008 min
Response: 3509341
Conc: 5.63 ng/ml



#5 AR-1016-3


R.T.: 6.642 min
Delta R.T.: -0.012 min
Response: 9322346
Conc: 20.50 ng/ml

Instrument :
ECD_Q
ClientSampleId :
CLI-2022-5A




#5 AR-1016-3

R.T.: 5.769 min
Delta R.T.: -0.011 min
Response: 1669026
Conc: 6.11 ng/ml




#6 AR-1016-4

R.T.: 6.778 min
Delta R.T.: 0.016 min
Response: 1672768
Conc: 4.52 ng/ml



#6 AR-1016-4


R.T.: 5.815 min
Delta R.T.: -0.012 min
Response: 1272799
Conc: 5.04 ng/ml



#7 AR-1016-5


R.T.: 7.068 min
Delta R.T.: -0.007 min
Response: 742700
Conc: 2.54 ng/ml

Instrument :
ECD_Q
ClientSampleId :
CLI-2022-5A




#7 AR-1016-5

R.T.: 6.048 min
Delta R.T.: -0.011 min
Response: 168061
Conc: 0.57 ng/ml




#8 AR-1221-1

R.T.: 5.498 min
Delta R.T.: 0.009 min
Response: 614950
Conc: 4.31 ng/ml



#8 AR-1221-1


R.T.: 4.582 min
Delta R.T.: 0.023 min
Response: 77647
Conc: 0.69 ng/ml



#9 AR-1221-2


R.T.: 5.598 min
Delta R.T.: 0.009 min
Response: 379932
Conc: 3.46 ng/ml

Instrument :
ECD_Q
ClientSampleId :
CLI-2022-5A




#9 AR-1221-2

R.T.: 4.652 min
Delta R.T.: -0.007 min
Response: 213332
Conc: 2.44 ng/ml




#10 AR-1221-3

R.T.: 5.669 min
Delta R.T.: -0.007 min
Response: 617925
Conc: 1.90 ng/ml



#10 AR-1221-3


R.T.: 4.748 min
Delta R.T.: -0.002 min
Response: 46938
Conc: 0.17 ng/ml



#11 AR-1232-1


R.T.: 5.669 min
Delta R.T.: -0.007 min
Response: 617925
Conc: 2.30 ng/ml

Instrument :
ECD_Q
ClientSampleId :
CLI-2022-5A




#11 AR-1232-1

R.T.: 4.748 min
Delta R.T.: 0.000 min
Response: 46938
Conc: 0.20 ng/ml




#12 AR-1232-2

R.T.: 6.281 min
Delta R.T.: 0.016 min
Response: 8383631
Conc: 56.55 ng/ml



#12 AR-1232-2


R.T.: 5.581 min
Delta R.T.: -0.007 min
Response: 3509341
Conc: 14.43 ng/ml



#13 AR-1232-3


R.T.: 6.563 min
Delta R.T.: -0.025 min
Response: 5269495
Conc: 18.96 ng/ml

Instrument :
ECD_Q
ClientSampleId :
CLI-2022-5A




#13 AR-1232-3

R.T.: 5.769 min
Delta R.T.: -0.011 min
Response: 1669026
Conc: 16.07 ng/ml




#14 AR-1232-4

R.T.: 6.778 min
Delta R.T.: 0.015 min
Response: 1672768
Conc: 11.31 ng/ml



#14 AR-1232-4


R.T.: 5.900 min
Delta R.T.: 0.028 min
Response: 927634
Conc: 8.94 ng/ml



#15 AR-1232-5


R.T.: 6.870 min
Delta R.T.: 0.016 min
Response: 1073757
Conc: 11.36 ng/ml

Instrument :
ECD_Q
ClientSampleId :
CLI-2022-5A




#15 AR-1232-5

R.T.: 6.048 min
Delta R.T.: -0.010 min
Response: 168061
Conc: 1.53 ng/ml




#16 AR-1242-1

R.T.: 6.563 min
Delta R.T.: -0.001 min
Response: 5269495
Conc: 19.53 ng/ml



#16 AR-1242-1


R.T.: 5.581 min
Delta R.T.: 0.013 min
Response: 3509341
Conc: 15.16 ng/ml



#17 AR-1242-2


R.T.: 6.563 min
Delta R.T.: -0.027 min
Response: 5269495
Conc: 10.07 ng/ml

Instrument :
ECD_Q
ClientSampleId :
CLI-2022-5A




#17 AR-1242-2

R.T.: 5.581 min
Delta R.T.: -0.008 min
Response: 3509341
Conc: 7.45 ng/ml




#18 AR-1242-3

R.T.: 6.642 min
Delta R.T.: -0.013 min
Response: 9322346
Conc: 25.78 ng/ml



#18 AR-1242-3


R.T.: 5.769 min
Delta R.T.: -0.012 min
Response: 1669026
Conc: 8.04 ng/ml



#19 AR-1242-4


R.T.: 6.778 min
Delta R.T.: 0.015 min
Response: 1672768
Conc: 5.73 ng/ml

Instrument :
ECD_Q
ClientSampleId :
CLI-2022-5A




#19 AR-1242-4

R.T.: 5.900 min
Delta R.T.: 0.027 min
Response: 927634
Conc: 4.16 ng/ml




#20 AR-1242-5

R.T.: 7.543 min
Delta R.T.: -0.003 min
Response: 750709
Conc: 2.83 ng/ml



#20 AR-1242-5


R.T.: 6.434 min
Delta R.T.: -0.007 min
Response: 674098
Conc: 2.50 ng/ml



#21 AR-1248-1


R.T.: 6.563 min
Delta R.T.: 0.000 min
Response: 5269495
Conc: 25.17 ng/ml

Instrument :
ECD_Q
ClientSampleId :
CLI-2022-5A




#21 AR-1248-1

R.T.: 5.581 min
Delta R.T.: 0.013 min
Response: 3509341
Conc: 20.30 ng/ml




#22 AR-1248-2

R.T.: 6.870 min
Delta R.T.: 0.016 min
Response: 1073757
Conc: 3.35 ng/ml



#22 AR-1248-2


R.T.: 5.815 min
Delta R.T.: -0.012 min
Response: 1272799
Conc: 3.99 ng/ml



#23 AR-1248-3


R.T.: 7.068 min
Delta R.T.: -0.006 min
Response: 742700
Conc: 1.96 ng/ml

Instrument :
ECD_Q
ClientSampleId :
CLI-2022-5A




#23 AR-1248-3

R.T.: 5.900 min
Delta R.T.: 0.027 min
Response: 927634
Conc: 2.83 ng/ml




#24 AR-1248-4

R.T.: 7.502 min
Delta R.T.: -0.003 min
Response: 241516
Conc: 0.57 ng/ml



#24 AR-1248-4


R.T.: 6.048 min
Delta R.T.: -0.010 min
Response: 168061
Conc: 0.45 ng/ml



#25 AR-1248-5


R.T.: 7.543 min
Delta R.T.: -0.003 min
Response: 750709
Conc: 1.63 ng/ml

Instrument :
ECD_Q
ClientSampleId :
CLI-2022-5A




#25 AR-1248-5

R.T.: 6.476 min
Delta R.T.: -0.010 min
Response: 227581
Conc: 0.62 ng/ml




#26 AR-1254-1

R.T.: 7.469 min
Delta R.T.: -0.002 min
Response: 248742
Conc: 0.61 ng/ml



#26 AR-1254-1


R.T.: 6.434 min
Delta R.T.: -0.005 min
Response: 674098
Conc: 1.09 ng/ml



#27 AR-1254-2


R.T.: 7.702 min
Delta R.T.: 0.001 min
Response: 1206676
Conc: 1.86 ng/ml

Instrument :
ECD_Q
ClientSampleId :
CLI-2022-5A




#27 AR-1254-2

R.T.: 6.592 min
Delta R.T.: -0.004 min
Response: 370446
Conc: 0.68 ng/ml




#28 AR-1254-3

R.T.: 8.083 min
Delta R.T.: -0.002 min
Response: 1221031
Conc: 1.56 ng/ml



#28 AR-1254-3


R.T.: 7.020 min
Delta R.T.: 0.000 min
Response: 345629
Conc: 0.40 ng/ml



#29 AR-1254-4


R.T.: 8.374 min
Delta R.T.: -0.008 min
Response: 1109263
Conc: 1.96 ng/ml

Instrument :
ECD_Q
ClientSampleId :
CLI-2022-5A




#29 AR-1254-4

R.T.: 7.257 min
Delta R.T.: -0.004 min
Response: 250185
Conc: 0.40 ng/ml




#30 AR-1254-5

R.T.: 8.810 min
Delta R.T.: 0.000 min
Response: 3599207
Conc: 5.70 ng/ml



#30 AR-1254-5


R.T.: 7.682 min
Delta R.T.: -0.007 min
Response: 1030031
Conc: 1.36 ng/ml



#31 AR-1260-1


R.T.: 8.246 min
Delta R.T.: -0.005 min
Response: 2648814
Conc: 5.51 ng/ml

Instrument :
ECD_Q
ClientSampleId :
CLI-2022-5A




#31 AR-1260-1

R.T.: 7.150 min
Delta R.T.: -0.006 min
Response: 627892
Conc: 1.15 ng/ml




#32 AR-1260-2

R.T.: 8.513 min
Delta R.T.: -0.006 min
Response: 3860444
Conc: 6.55 ng/ml



#32 AR-1260-2


R.T.: 7.345 min
Delta R.T.: -0.007 min
Response: 870658
Conc: 1.31 ng/ml



#33 AR-1260-3


R.T.: 8.881 min
Delta R.T.: -0.005 min
Response: 973049
Conc: 1.95 ng/ml

Instrument :
ECD_Q
ClientSampleId :
CLI-2022-5A




#33 AR-1260-3

R.T.: 7.530 min
Delta R.T.: 0.020 min
Response: 333997
Conc: 0.54 ng/ml




#34 AR-1260-4

R.T.: 9.127 min
Delta R.T.: 0.000 min
Response: 1763460
Conc: 3.03 ng/ml



#34 AR-1260-4


R.T.: 7.986 min
Delta R.T.: -0.007 min
Response: 633509
Conc: 1.25 ng/ml



#35 AR-1260-5


R.T.: 9.472 min
Delta R.T.: -0.002 min
Response: 8666271
Conc: 6.59 ng/ml

Instrument :
ECD_Q
ClientSampleId :
CLI-2022-5A




#35 AR-1260-5

R.T.: 8.232 min
Delta R.T.: -0.007 min
Response: 1490571
Conc: 1.25 ng/ml




#36 AR-1262-1

R.T.: 8.881 min
Delta R.T.: -0.002 min
Response: 973049
Conc: 1.36 ng/ml



#36 AR-1262-1


R.T.: 7.682 min
Delta R.T.: -0.006 min
Response: 1030031
Conc: 2.83 ng/ml



#37 AR-1262-2


R.T.: 9.472 min
Delta R.T.: 0.000 min
Response: 8666271
Conc: 5.79 ng/ml

Instrument :
ECD_Q
ClientSampleId :
CLI-2022-5A




#37 AR-1262-2

R.T.: 8.232 min
Delta R.T.: -0.006 min
Response: 1490571
Conc: 1.13 ng/ml




#38 AR-1262-3

R.T.: 9.799 min
Delta R.T.: -0.003 min
Response: 434005
Conc: 0.56 ng/ml



#38 AR-1262-3


R.T.: 8.520 min
Delta R.T.: -0.004 min
Response: 853260
Conc: 1.72 ng/ml



#39 AR-1262-4


R.T.: 9.914 min
Delta R.T.: 0.009 min
Response: 756715
Conc: 1.14 ng/ml

Instrument :
ECD_Q
ClientSampleId :
CLI-2022-5A




#39 AR-1262-4

R.T.: 8.582 min
Delta R.T.: -0.007 min
Response: 1108160
Conc: 1.20 ng/ml




#40 AR-1262-5

R.T.: 10.611 min
Delta R.T.: -0.005 min
Response: 2410789
Conc: 3.52 ng/ml



#40 AR-1262-5


R.T.: 9.095 min
Delta R.T.: -0.001 min
Response: 378376
Conc: 0.95 ng/ml



#41 AR-1268-1


R.T.: 9.799 min
Delta R.T.: -0.004 min
Response: 434005
Conc: 0.23 ng/ml

Instrument :
ECD_Q
ClientSampleId :
CLI-2022-5A




#41 AR-1268-1

R.T.: 8.520 min
Delta R.T.: -0.004 min
Response: 853260
Conc: 0.58 ng/ml




#42 AR-1268-2

R.T.: 9.914 min
Delta R.T.: 0.008 min
Response: 756715
Conc: 0.38 ng/ml



#42 AR-1268-2


R.T.: 8.582 min
Delta R.T.: -0.007 min
Response: 1108160
Conc: 0.84 ng/ml



#43 AR-1268-3


R.T.: 10.166 min
Delta R.T.: 0.019 min
Response: 1352145
Conc: 0.79 ng/ml

Instrument :
ECD_Q
ClientSampleId :
CLI-2022-5A




#43 AR-1268-3

R.T.: 8.789 min
Delta R.T.: -0.010 min
Response: 590959
Conc: 0.54 ng/ml




#44 AR-1268-4

R.T.: 10.611 min
Delta R.T.: -0.006 min
Response: 2410789
Conc: 3.23 ng/ml



#44 AR-1268-4


R.T.: 9.095 min
Delta R.T.: 0.000 min
Response: 378376
Conc: 0.86 ng/ml



#45 AR-1268-5

R.T.: 11.056 min
Delta R.T.: -0.006 min
Response: 612338
Conc: 0.11 ng/ml

Instrument :
ECD_Q
ClientSampleId :
CLI-2022-5A



#45 AR-1268-5

R.T.: 9.398 min
Delta R.T.: 0.000 min
Response: 216810
Conc: 0.07 ng/ml