

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ012621\
 Data File : PQ051916.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Jan 2021 01:39
 Operator : DD\AJ
 Sample : AR1262ICC1600
 Misc :
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1262501

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 27 06:08:43 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ012621CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jan 27 05:59:30 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	5.231	4.248	1867.7E6	627.5E6	72.635	73.289
2) SA Decachlor...	11.420	9.595	3189.8E6	1160.0E6	150.006	151.787
Target Compounds						
36) L8 AR-1262-1	8.871	7.628	2370.4E6	472.0E6	1519.223	1540.897
37) L8 AR-1262-2	9.459	8.179	4504.3E6	1764.0E6	1576.970	1541.575
38) L8 AR-1262-3	9.791	8.467	2906.4E6	680.0E6	1542.440	1531.495
39) L8 AR-1262-4	9.891	8.531	2192.7E6	1290.8E6	1517.146	1556.004
40) L8 AR-1262-5	10.608	9.033	1494.5E6	559.2E6	1543.589	1562.370

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ012621\
Data File : PQ051916.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Jan 2021 01:39
Operator : DD\AJ
Sample : AR1262ICC1600
Misc :
ALS Vial : 37 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1262501

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 27 06:08:43 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ012621CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jan 27 05:59:30 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

