

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ012722\
 Data File : PQ056018.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Jan 2022 15:40
 Operator : AJ\MA
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1248CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 01/28/2022
 Supervised By :Ankita Jodhani 01/28/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 27 16:40:09 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ011922.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jan 20 13:18:56 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	5.236	4.294	738.3E6	657.7E6	54.511	58.492
2)	SA Decachlor...	11.409	9.651	790.5E6	453.9E6	54.152	56.910
Target Compounds							
21)	L5 AR-1248-1	6.565	5.564	95612692	95836538	456.609	554.417
22)	L5 AR-1248-2	6.851	5.820	171.0E6	179.0E6	533.935	561.237
23)	L5 AR-1248-3	7.070	5.867	197.9E6	183.5E6	523.017	560.554
24)	L5 AR-1248-4	7.503	6.054	212.2E6	198.5E6	501.560	526.299
25)	L5 AR-1248-5	7.544	6.482	230.3E6	208.5E6	500.127	567.095m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ012722\
Data File : PQ056018.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Jan 2022 15:40
Operator : AJ\MA
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1248CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 01/28/2022
Supervised By :Ankita Jodhani 01/28/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 27 16:40:09 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ011922.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jan 20 13:18:56 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm

