

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ012722\
 Data File : PQ056041.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Jan 2022 22:21
 Operator : AJ\MA
 Sample : PB142309BSD
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 PB142309BSD

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 01/28/2022
 Supervised By :Ankita Jodhani 01/28/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 28 04:54:55 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ011922.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jan 20 13:18:56 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	5.235	4.295	307.0E6	240.3E6	22.666	21.369
2) SA Decachlor...	11.411	9.651	380.5E6	191.7E6	26.066	24.041m
Target Compounds						
3) L1 AR-1016-1	6.563	5.562	154.0E6	155.1E6	440.271	505.082
4) L1 AR-1016-2	6.587	5.583	331.1E6	318.7E6	487.028	511.675
5) L1 AR-1016-3	6.653	5.776	228.4E6	136.4E6	502.368	499.650
6) L1 AR-1016-4	6.762	5.821	180.8E6	120.4E6	488.146	477.044
7) L1 AR-1016-5	7.071	6.052	140.9E6	140.2E6	481.144	478.899
31) L7 AR-1260-1	8.246	7.148	253.7E6	280.4E6	527.854	512.051
32) L7 AR-1260-2	8.511	7.344	312.8E6	346.8E6	531.041	522.464
33) L7 AR-1260-3	8.877	7.501	235.9E6	320.5E6	471.829	520.655
34) L7 AR-1260-4	9.120	7.983	283.5E6	221.0E6	486.633	437.581
35) L7 AR-1260-5	9.466	8.230	645.5E6	531.8E6	491.148	447.736

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ012722\
Data File : PQ056041.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Jan 2022 22:21
Operator : AJ\MA
Sample : PB142309BSD
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
PB142309BSD

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 01/28/2022
Supervised By :Ankita Jodhani 01/28/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 28 04:54:55 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ011922.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jan 20 13:18:56 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

