

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ012821\
 Data File : PQ051958.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Jan 2021 18:48
 Operator : DD\AJ
 Sample : M1264-06
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 S1-06-E

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 29 02:43:08 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ012821.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jan 29 02:22:28 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	5.232	4.249	767.9E6	240.1E6	33.786	32.972
2) SA Decachlor...	11.422	9.596	504.9E6	178.3E6	25.615	25.858
Target Compounds						
26) L6 AR-1254-1	7.464	6.382	357.8E6	118.2E6	398.438	285.927 #
27) L6 AR-1254-2	7.692	6.540	652.0E6	199.9E6	461.858	562.152
28) L6 AR-1254-3	8.077	6.962	923.6E6	397.6E6	602.041	676.579
29) L6 AR-1254-4	8.371	7.200	875.9E6	305.2E6	826.786	890.662
30) L6 AR-1254-5	8.801	7.630	1813.9E6	1234.9E6	1533.517	2462.007 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ012821\
Data File : PQ051958.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Jan 2021 18:48
Operator : DD\AJ
Sample : M1264-06
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampled :
S1-06-E

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 29 02:43:08 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ012821.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jan 29 02:22:28 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 5.232 min
Delta R.T.: -0.001 min
Response: 767873458
Conc: 33.79 ng/ml

Instrument :
ECD_Q
ClientSampleId :
S1-06-E




#1 Tetrachloro-m-xylene

R.T.: 4.249 min
Delta R.T.: -0.002 min
Response: 240149424
Conc: 32.97 ng/ml




#2 Decachlorobiphenyl

R.T.: 11.422 min
Delta R.T.: -0.002 min
Response: 504873725
Conc: 25.62 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.596 min
Delta R.T.: 0.000 min
Response: 178333671
Conc: 25.86 ng/ml



#26 AR-1254-1


R.T.: 7.464 min
Delta R.T.: -0.002 min
Response: 357803219
Conc: 398.44 ng/ml

Instrument :
ECD_Q
ClientSampleId :
S1-06-E




#26 AR-1254-1

R.T.: 6.382 min
Delta R.T.: 0.000 min
Response: 118173309
Conc: 285.93 ng/ml




#27 AR-1254-2

R.T.: 7.692 min
Delta R.T.: -0.001 min
Response: 652029899
Conc: 461.86 ng/ml



#27 AR-1254-2


R.T.: 6.540 min
Delta R.T.: 0.000 min
Response: 199853892
Conc: 562.15 ng/ml



#28 AR-1254-3


R.T.: 8.077 min
Delta R.T.: 0.000 min
Response: 923610602
Conc: 602.04 ng/ml

Instrument :
ECD_Q
ClientSampleId :
S1-06-E




#28 AR-1254-3

R.T.: 6.962 min
Delta R.T.: 0.000 min
Response: 397607832
Conc: 676.58 ng/ml




#29 AR-1254-4

R.T.: 8.371 min
Delta R.T.: 0.000 min
Response: 875913652
Conc: 826.79 ng/ml



#29 AR-1254-4


R.T.: 7.200 min
Delta R.T.: -0.001 min
Response: 305200458
Conc: 890.66 ng/ml



#30 AR-1254-5

R.T.: 8.801 min
Delta R.T.: 0.000 min
Response: 1813909748
Conc: 1533.52 ng/ml

Instrument :
ECD_Q
ClientSampleId :
S1-06-E



#30 AR-1254-5

R.T.: 7.630 min
Delta R.T.: 0.000 min
Response: 1234895996
Conc: 2462.01 ng/ml