

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ013020\
 Data File : PQ046275.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Jan 2020 15:34
 Operator : AJ\MA
 Sample : AR1221ICC500
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1221ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 30 16:43:37 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ013020.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jan 30 16:43:24 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	5.210	4.346	342.8E6	168.2E6	50.000	50.000
2) SA Decachlor...	11.358	9.780	280.1E6	268.9E6	50.000	50.000
Target Compounds						
8) L2 AR-1221-1	5.454	4.609	39337467	17599903	500.000	500.000
9) L2 AR-1221-2	5.553	4.712	30415380	13116891	500.000	500.000
10) L2 AR-1221-3	5.640	4.803	94508829	43880745	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ013020\
Data File : PQ046275.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Jan 2020 15:34
Operator : AJ\MA
Sample : AR1221ICC500
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1221ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 30 16:43:37 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ013020.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jan 30 16:43:24 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 5.210 min
Delta R.T.: 0.000 min
Response: 342789016
Conc: 50.00 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1221ICC500




#1 Tetrachloro-m-xylene

R.T.: 4.346 min
Delta R.T.: 0.000 min
Response: 168219996
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 11.358 min
Delta R.T.: 0.000 min
Response: 280103992
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.780 min
Delta R.T.: 0.000 min
Response: 268928336
Conc: 50.00 ng/ml



#8 AR-1221-1


R.T.: 5.454 min
Delta R.T.: 0.000 min
Response: 39337467
Conc: 500.00 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1221ICC500




#8 AR-1221-1

R.T.: 4.609 min
Delta R.T.: 0.000 min
Response: 17599903
Conc: 500.00 ng/ml




#9 AR-1221-2

R.T.: 5.553 min
Delta R.T.: 0.000 min
Response: 30415380
Conc: 500.00 ng/ml



#9 AR-1221-2


R.T.: 4.712 min
Delta R.T.: 0.000 min
Response: 13116891
Conc: 500.00 ng/ml



#10 AR-1221-3

R.T.: 5.640 min
Delta R.T.: 0.000 min
Response: 94508829
Conc: 500.00 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1221ICC500



#10 AR-1221-3

R.T.: 4.803 min
Delta R.T.: 0.000 min
Response: 43880745
Conc: 500.00 ng/ml