

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ020723\
 Data File : PQ060364.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Feb 2023 10:37
 Operator : YP\AJ
 Sample : AR1262ICC1600
 Misc :
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR12625015

Manual Integrations
 APPROVED

Reviewed By :Ankita Jodhani 02/09/2023
 Supervised By :Sohil Jodhani 02/09/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 08 11:05:13 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ020723CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Feb 08 11:03:36 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.440	2.799	1622.5E6	907.2E6	70.576	72.433
2) SA Decachlor...	8.571	7.533	1911.9E6	1521.3E6	136.274	138.950m
Target Compounds						
36) L8 AR-1262-1	6.667	5.712	1686.7E6	1107.5E6	1338.353	1391.279
37) L8 AR-1262-2	7.196	6.207	3071.5E6	2099.3E6	1437.274	1438.723
38) L8 AR-1262-3	7.473	6.487	1954.2E6	814.0E6	1399.250	1390.667
39) L8 AR-1262-4	7.546	6.549	1478.7E6	1567.9E6	1373.122	1413.698
40) L8 AR-1262-5	8.052	7.045	973.2E6	750.8E6	1390.974m	1394.911

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ020723\
Data File : PQ060364.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Feb 2023 10:37
Operator : YP\AJ
Sample : AR1262ICC1600
Misc :
ALS Vial : 37 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR12625015

Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 02/09/2023
Supervised By :Sohil Jodhani 02/09/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 08 11:05:13 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ020723CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Feb 08 11:03:36 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

