Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ020821\
Data File : PQ052078.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Feb 2021 18:23
Operator : DD\AJ
Sample : M1317-05
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
R-C

Manual Integrations APPROVED
Ankita
2/9/2021 10:48:24 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 09 06:34:01 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ020821.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Feb 09 06:20:03 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	5.238	4.261	326.3E6	117.3E6	13.107	13.605
2)	SA Decachlor...	11.423	9.597	306.2E6	107.1E6	17.747	15.704
	Target Compounds						
31)	L7 AR-1260-1	8.247	7.109	35087359	17249757	35.740m	43.383
32)	L7 AR-1260-2	8.508	7.302	91419927	22315211	78.542m	45.169 #
33)	L7 AR-1260-3	8.875	7.457	28636361	14827687	33.802m	32.321
34)	L7 AR-1260-4	9.116	7.939	41824543	10975409	42.662m	28.693 #
35)	L7 AR-1260-5	9.463	8.183	75694088	28935908	38.503	31.087

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ020821\
Data File : PQ052078.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Feb 2021 18:23
Operator : DD\AJ
Sample : M1317-05
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
R-C

Manual Integrations APPROVED
Ankita
2/9/2021 10:48:24 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 09 06:34:01 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ020821.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Feb 09 06:20:03 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm