

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ020824\
 Data File : PQ065342.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Feb 2024 23:20
 Operator : YP\AJ
 Sample : P1412-01
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 TP2

Manual Integrations
 APPROVED

Reviewed By :Ankita Jodhani 02/09/2024
 Supervised By :Sohil Jodhani 02/09/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 09 00:49:04 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ011724.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jan 18 00:09:27 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.465	2.741	73885280	154.3E6	19.833	23.105
2) SA Decachlor...	8.771	7.513	54110954	126.5E6	17.509	17.402
Target Compounds						
31) L7 AR-1260-1	6.176	5.153	8445102	24374998	47.957	60.566m#
32) L7 AR-1260-2	6.442	5.343	10960507	25642755	47.980	52.434m
33) L7 AR-1260-3	6.804	5.487	15586263	26435997	97.075	55.268 #
34) L7 AR-1260-4	7.025	5.948	14907332	35407752	83.824	90.996
35) L7 AR-1260-5	7.344	6.194	33967742	88392749	92.523	101.281

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ020824\
Data File : PQ065342.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Feb 2024 23:20
Operator : YP\AJ
Sample : P1412-01
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
TP2

Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 02/09/2024
Supervised By :Sohil Jodhani 02/09/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 09 00:49:04 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ011724.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jan 18 00:09:27 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

