

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ021221\
 Data File : PQ052180.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Feb 2021 21:23
 Operator : DD\AJ
 Sample : AR1232ICC800
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 12 05:43:03 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ021221CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Feb 12 05:39:09 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	5.229	4.245	1079.8E6	383.1E6	39.692	40.945
2) SA Decachlor...	11.419	9.592	1742.0E6	703.0E6	79.863	80.805
Target Compounds						
11) L3 AR-1232-1	5.665	4.692	373.7E6	140.1E6	773.538	773.888
12) L3 AR-1232-2	6.258	5.527	201.1E6	153.6E6	788.584	790.181
13) L3 AR-1232-3	6.577	5.720	411.7E6	77647245	786.933	787.041
14) L3 AR-1232-4	6.752	5.815	202.7E6	64715801	756.959	762.411
15) L3 AR-1232-5	6.847	6.000	139.8E6	69580048	739.682	781.692



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ021221\
Data File : PQ052180.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Feb 2021 21:23
Operator : DD\AJ
Sample : AR1232ICC800
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 12 05:43:03 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ021221CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Feb 12 05:39:09 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 5.229 min
Delta R.T.: 0.000 min
Response: 1079785953
Conc: 39.69 ng/ml

Instrument :
ECD_Q
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.245 min
Delta R.T.: -0.001 min
Response: 383107160
Conc: 40.95 ng/ml




#2 Decachlorobiphenyl

R.T.: 11.419 min
Delta R.T.: -0.001 min
Response: 1741975086
Conc: 79.86 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.592 min
Delta R.T.: 0.000 min
Response: 702991408
Conc: 80.81 ng/ml



#11 AR-1232-1


R.T.: 5.665 min
Delta R.T.: 0.000 min
Response: 373701172
Conc: 773.54 ng/ml

Instrument :
ECD_Q
ClientSampleId :




#11 AR-1232-1

R.T.: 4.692 min
Delta R.T.: 0.000 min
Response: 140059836
Conc: 773.89 ng/ml




#12 AR-1232-2

R.T.: 6.258 min
Delta R.T.: 0.000 min
Response: 201094249
Conc: 788.58 ng/ml



#12 AR-1232-2


R.T.: 5.527 min
Delta R.T.: 0.000 min
Response: 153609903
Conc: 790.18 ng/ml



#13 AR-1232-3


R.T.: 6.577 min
Delta R.T.: 0.000 min
Response: 411712070
Conc: 786.93 ng/ml

Instrument :
ECD_Q
ClientSampleId :




#13 AR-1232-3

R.T.: 5.720 min
Delta R.T.: 0.000 min
Response: 77647245
Conc: 787.04 ng/ml




#14 AR-1232-4

R.T.: 6.752 min
Delta R.T.: 0.000 min
Response: 202723202
Conc: 756.96 ng/ml



#14 AR-1232-4


R.T.: 5.815 min
Delta R.T.: 0.000 min
Response: 64715801
Conc: 762.41 ng/ml



#15 AR-1232-5

R.T.: 6.847 min
Delta R.T.: 0.000 min
Response: 139847030
Conc: 739.68 ng/ml

Instrument :
ECD_Q
ClientSampleId :



#15 AR-1232-5

R.T.: 6.000 min
Delta R.T.: 0.000 min
Response: 69580048
Conc: 781.69 ng/ml