

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ021224\
 Data File : PQ065426.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Feb 2024 03:41
 Operator : YP\AJ
 Sample : AR1268ICC750
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1268ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 13 04:18:31 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ021224.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Feb 13 04:12:50 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.475	2.741	275.8E6	564.3E6	74.268	74.436
2) SA Decachlor...	8.797	7.515	376.1E6	965.3E6	74.685	74.161
Target Compounds						
41) L9 AR-1268-1	7.647	6.474	320.5E6	865.3E6	746.488	742.865
42) L9 AR-1268-2	7.727	6.537	291.8E6	800.8E6	747.613	742.982
43) L9 AR-1268-3	7.912	6.736	248.3E6	692.5E6	747.565	741.951
44) L9 AR-1268-4	8.248	7.029	110.5E6	299.7E6	747.890	742.895
45) L9 AR-1268-5	8.554	7.298	768.8E6	2062.5E6	748.573	743.249



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ021224\
Data File : PQ065426.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Feb 2024 03:41
Operator : YP\AJ
Sample : AR1268ICC750
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1268ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 13 04:18:31 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ021224.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Feb 13 04:12:50 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.475 min
Delta R.T.: 0.000 min
Response: 275793247
Conc: 74.27 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1268ICC750




#1 Tetrachloro-m-xylene

R.T.: 2.741 min
Delta R.T.: 0.000 min
Response: 564327826
Conc: 74.44 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.797 min
Delta R.T.: 0.000 min
Response: 376085729
Conc: 74.69 ng/ml



#2 Decachlorobiphenyl


R.T.: 7.515 min
Delta R.T.: 0.000 min
Response: 965336099
Conc: 74.16 ng/ml



#41 AR-1268-1


R.T.: 7.647 min
Delta R.T.: 0.000 min
Response: 320527212
Conc: 746.49 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1268ICC750




#41 AR-1268-1

R.T.: 6.474 min
Delta R.T.: 0.000 min
Response: 865267346
Conc: 742.87 ng/ml




#42 AR-1268-2

R.T.: 7.727 min
Delta R.T.: 0.000 min
Response: 291751772
Conc: 747.61 ng/ml



#42 AR-1268-2


R.T.: 6.537 min
Delta R.T.: 0.000 min
Response: 800829454
Conc: 742.98 ng/ml



#43 AR-1268-3


R.T.: 7.912 min
Delta R.T.: 0.000 min
Response: 248324667
Conc: 747.57 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1268ICC750




#43 AR-1268-3

R.T.: 6.736 min
Delta R.T.: 0.000 min
Response: 692488759
Conc: 741.95 ng/ml




#44 AR-1268-4

R.T.: 8.248 min
Delta R.T.: 0.000 min
Response: 110531622
Conc: 747.89 ng/ml



#44 AR-1268-4


R.T.: 7.029 min
Delta R.T.: 0.000 min
Response: 299727847
Conc: 742.90 ng/ml



#45 AR-1268-5

R.T.: 8.554 min
Delta R.T.: 0.000 min
Response: 768753798
Conc: 748.57 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1268ICC750



#45 AR-1268-5

R.T.: 7.298 min
Delta R.T.: 0.000 min
Response: 2062516057
Conc: 743.25 ng/ml