

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ021622\
 Data File : PQ056290.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Feb 2022 13:47
 Operator : YP\AJ
 Sample : N1506-16DL 4X
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 BGZG4DL

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 02/17/2022
 Supervised By :Ankita Jodhani 02/17/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 17 02:03:46 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ021422CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Feb 15 04:32:21 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.683	3.854	70467184	47569730	4.760	3.985
2) SA Decachlor...	10.589	8.841	141.7E6	47918046	7.930	4.742m#
Target Compounds						
26) L6 AR-1254-1	6.719	5.719	1502.7E6	2550.5E6	3636.027	4112.764
27) L6 AR-1254-2	6.937	5.865	2557.8E6	2263.2E6	3830.061	4278.934
28) L6 AR-1254-3	7.308	6.266	3648.3E6	4156.4E6	4469.497	4804.713
29) L6 AR-1254-4	7.595	6.492	2476.5E6	2309.6E6	4289.036	3700.151
30) L6 AR-1254-5	8.015	6.905	4496.7E6	4396.8E6	5785.619	5626.179

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ021622\
Data File : PQ056290.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Feb 2022 13:47
Operator : YP\AJ
Sample : N1506-16DL 4X
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
BGZG4DL

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/17/2022
Supervised By :Ankita Jodhani 02/17/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 17 02:03:46 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ021422CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Feb 15 04:32:21 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

