

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ022219\
 Data File : PQ037477.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Feb 2019 17:35
 Operator : SM\SJ
 Sample : AR1242ICC500
 Misc :
 ALS Vial : 88 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1242ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 23 03:21:18 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ022219.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Feb 23 03:20:55 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.244	3.659	270.8E6	144.2E6	50.000	50.000
2)	SA Decachlor...	9.732	8.560	211.0E6	93339923	50.000	50.000
Target Compounds							
16)	L4 AR-1242-1	5.392	4.722	73493493	40843099	500.000	500.000
17)	L4 AR-1242-2	5.412	4.738	113.0E6	61450717	500.000	500.000
18)	L4 AR-1242-3	5.473	4.913	66241725	30975654	500.000	500.000
19)	L4 AR-1242-4	5.573	4.993	54532200	29683725	500.000	500.000
20)	L4 AR-1242-5	6.293	5.507	60038397	35975241	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ022219\
Data File : PQ037477.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Feb 2019 17:35
Operator : SM\SJ
Sample : AR1242ICC500
Misc :
ALS Vial : 88 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1242ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 23 03:21:18 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ022219.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Feb 23 03:20:55 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

