

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ022219\
 Data File : PQ037495.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Feb 2019 07:17
 Operator : SM\SJ
 Sample : K1550-03
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 SS022019

Manual Integrations
 APPROVED

Sohil
 2/25/2019 5:26:17 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 24 23:54:30 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ022219.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Feb 23 04:01:25 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

1) SA Tetrachlo...	4.250	3.666	123.4E6	51898579	19.690	15.533
2) SA Decachlor...	9.730	8.556	94913645	39514248	18.386	17.133m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ022219\
Data File : PQ037495.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Feb 2019 07:17
Operator : SM\SJ
Sample : K1550-03
Misc :
ALS Vial : 33 Sample Multiplier: 1



Instrument :
ECD_Q
ClientSampleId :
SS022019


Manual Integrations
APPROVED

Sohil
2/25/2019 5:26:17 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 24 23:54:30 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ022219.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Feb 23 04:01:25 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.250 min
Delta R.T.: 0.004 min
Response: 123361375
Conc: 19.69 ng/ml

Instrument :
ECD_Q
ClientSampleId :
SS022019


Manual Integrations
APPROVED

Sohil
2/25/2019 5:26:17 PM




#1 Tetrachloro-m-xylene

R.T.: 3.666 min
Delta R.T.: 0.006 min
Response: 51898579
Conc: 15.53 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.730 min
Delta R.T.: -0.003 min
Response: 94913645
Conc: 18.39 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.556 min
Delta R.T.: -0.004 min
Response: 39514248
Conc: 17.13 ng/ml m