

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ022423\
 Data File : PQ060471.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Feb 2023 02:14
 Operator : YP\AJ
 Sample : 01627-08
 Misc : AR1221 LOQ 50 PPB
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 LOQ-WATER-02-QT1-2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 25 08:38:21 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ022423.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Feb 24 16:56:07 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.439	2.799	500.4E6	255.4E6	21.747	22.699
2) SA Decachlor...	8.576	7.536	321.5E6	254.1E6	23.713	23.676
Target Compounds						
8) L2 AR-1221-1	3.637	2.999	17776786	7632160	74.207	67.902
9) L2 AR-1221-2	3.717	3.073	19472806	8217856	111.402	98.719
10) L2 AR-1221-3	3.786	3.143	42216700	17856690	80.698	69.113

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ022423\
Data File : PQ060471.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Feb 2023 02:14
Operator : YP\AJ
Sample : 01627-08
Misc : AR1221 LOQ 50 PPB
ALS Vial : 45 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
LOQ-WATER-02-QT1-2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 25 08:38:21 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ022423.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Feb 24 16:56:07 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm

