

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ022619\
 Data File : PQ037538.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Feb 2019 23:17
 Operator : SM\SJ
 Sample : AR1254ICC1600
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1254501

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 26 03:48:40 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ022619CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Feb 26 03:47:00 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.234	3.659	404.2E6	206.7E6	70.818	69.848
2) SA Decachlor...	9.705	8.551	808.3E6	368.9E6	158.131	153.970
Target Compounds						
26) L6 AR-1254-1	6.211	5.502	325.0E6	254.6E6	1567.691	1559.979
27) L6 AR-1254-2	6.422	5.647	524.4E6	218.8E6	1613.803	1574.651
28) L6 AR-1254-3	6.791	6.044	572.7E6	372.7E6	1736.539	1682.894
29) L6 AR-1254-4	7.073	6.270	446.3E6	272.2E6	1676.008	1627.327
30) L6 AR-1254-5	7.485	6.679	463.1E6	316.5E6	1782.834	1692.591

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ022619\
Data File : PQ037538.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Feb 2019 23:17
Operator : SM\SJ
Sample : AR1254ICC1600
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1254501

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 26 03:48:40 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ022619CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Feb 26 03:47:00 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

