

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ022720\
 Data File : PQ046629.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Feb 2020 12:57
 Operator : AJ\MA
 Sample : PB127117BL
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 28 03:12:55 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ022120.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Feb 21 13:47:31 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	5.197	4.329	377.4E6	195.6E6	24.838	25.724
2)	SA Decachlor...	11.347	9.761	293.3E6	146.1E6	36.035	21.094 #
Target Compounds							
7)	L1 AR-1016-5	6.994	0.000	14738046	0	39.290	N.D. #
9)	L2 AR-1221-2	5.550	4.693	4579485	2437041	42.044	48.587
10)	L2 AR-1221-3	5.550	0.000	4579485	0	13.674	N.D. #
11)	L3 AR-1232-1	5.550	0.000	4579485	0	16.957	N.D. #
23)	L5 AR-1248-3	6.994	0.000	14738046	0	35.690	N.D. #
28)	L6 AR-1254-3	8.052	0.000	2191971	0	3.023	N.D. #
43)	L9 AR-1268-3	0.000	8.875	0	18467779	N.D.	23.519 #
45)	L9 AR-1268-5	0.000	9.413	0	6185369	N.D.	2.446 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ022720\
Data File : PQ046629.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Feb 2020 12:57
Operator : AJ\MA
Sample : PB127117BL
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 28 03:12:55 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ022120.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Feb 21 13:47:31 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 5.197 min
Delta R.T.: 0.012 min
Response: 377395629
Conc: 24.84 ng/ml

Instrument :
ECD_Q
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.329 min
Delta R.T.: -0.005 min
Response: 195592916
Conc: 25.72 ng/ml




#2 Decachlorobiphenyl

R.T.: 11.347 min
Delta R.T.: 0.031 min
Response: 293345927
Conc: 36.04 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.761 min
Delta R.T.: 0.007 min
Response: 146114014
Conc: 21.09 ng/ml



#7 AR-1016-5


R.T.: 6.994 min
Delta R.T.: -0.020 min
Response: 14738046
Conc: 39.29 ng/ml

Instrument :
ECD_Q
ClientSampleId :




#7 AR-1016-5

R.T.: 0.000 min
Exp R.T. : 6.116 min
Response: 0
Conc: N.D.




#9 AR-1221-2

R.T.: 5.550 min
Delta R.T.: 0.021 min
Response: 4579485
Conc: 42.04 ng/ml



#9 AR-1221-2


R.T.: 4.693 min
Delta R.T.: -0.007 min
Response: 2437041
Conc: 48.59 ng/ml



#10 AR-1221-3


R.T.: 5.550 min
Delta R.T.: -0.066 min
Response: 4579485
Conc: 13.67 ng/ml

Instrument :
ECD_Q
ClientSampleId :




#10 AR-1221-3

R.T.: 0.000 min
Exp R.T. : 4.792 min
Response: 0
Conc: N.D.




#11 AR-1232-1

R.T.: 5.550 min
Delta R.T.: -0.066 min
Response: 4579485
Conc: 16.96 ng/ml



#11 AR-1232-1


R.T.: 0.000 min
Exp R.T. : 4.792 min
Response: 0
Conc: N.D.



#23 AR-1248-3


R.T.: 6.994 min
Delta R.T.: -0.021 min
Response: 14738046
Conc: 35.69 ng/ml

Instrument :
ECD_Q
ClientSampleId :




#23 AR-1248-3

R.T.: 0.000 min
Exp R.T. : 5.926 min
Response: 0
Conc: N.D.




#28 AR-1254-3

R.T.: 8.052 min
Delta R.T.: 0.028 min
Response: 2191971
Conc: 3.02 ng/ml



#28 AR-1254-3


R.T.: 0.000 min
Exp R.T. : 7.082 min
Response: 0
Conc: N.D.



#43 AR-1268-3


R.T.: 0.000 min
Exp R.T.: 10.070 min
Response: 0
Conc: N.D.

Instrument :
ECD_Q
ClientSampleId :




#43 AR-1268-3

R.T.: 8.875 min
Delta R.T.: 0.005 min
Response: 18467779
Conc: 23.52 ng/ml



#45 AR-1268-5

R.T.: 0.000 min
Exp R.T.: 10.960 min
Response: 0
Conc: N.D.



#45 AR-1268-5

R.T.: 9.413 min
Delta R.T.: -0.070 min
Response: 6185369
Conc: 2.45 ng/ml