

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ022723\
 Data File : PQ060537.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Feb 2023 21:43
 Operator : YP\AJ
 Sample : 01627-07
 Misc : AR1254 LOD 40 PPB
 ALS Vial : 40 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 27 23:38:47 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ022423.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Feb 24 16:56:07 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml



System Monitoring Compounds							
1)	SA Tetrachlo...	3.435	2.796	499.1E6	266.6E6	21.691	23.696
2)	SA Decachlor...	8.558	7.517	300.9E6	253.3E6	22.195	23.598
Target Compounds							
26)	L6 AR-1254-1	5.320	4.527	47217544	33798954	56.448	52.198
27)	L6 AR-1254-2	5.532	4.671	70699315	28915214	55.258	52.491
28)	L6 AR-1254-3	5.878	5.046	69836069	40857213	51.405	47.649m
29)	L6 AR-1254-4	6.160	5.267	53265251	27960845	53.706	49.939
30)	L6 AR-1254-5	6.570	5.664	56631961	40466876	53.947	49.928


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ022723\
Data File : PQ060537.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Feb 2023 21:43
Operator : YP\AJ
Sample : 01627-07
Misc : AR1254 LOD 40 PPB
ALS Vial : 40 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 27 23:38:47 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ022423.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Feb 24 16:56:07 2023
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.435 min
Delta R.T.: -0.005 min
Response: 499120382
Conc: 21.69 ng/ml




#1 Tetrachloro-m-xylene

R.T.: 2.796 min
Delta R.T.: -0.002 min
Response: 266575152
Conc: 23.70 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.558 min
Delta R.T.: -0.013 min
Response: 300949871
Conc: 22.20 ng/ml




#2 Decachlorobiphenyl

R.T.: 7.517 min
Delta R.T.: -0.017 min
Response: 253264496
Conc: 23.60 ng/ml




#26 AR-1254-1

R.T.: 5.320 min
Delta R.T.: -0.007 min
Response: 47217544
Conc: 56.45 ng/ml




#26 AR-1254-1

R.T.: 4.527 min
Delta R.T.: -0.005 min
Response: 33798954
Conc: 52.20 ng/ml




#27 AR-1254-2

R.T.: 5.532 min
Delta R.T.: -0.007 min
Response: 70699315
Conc: 55.26 ng/ml




#27 AR-1254-2

R.T.: 4.671 min
Delta R.T.: -0.005 min
Response: 28915214
Conc: 52.49 ng/ml




#28 AR-1254-3

R.T.: 5.878 min
Delta R.T.: -0.009 min
Response: 69836069
Conc: 51.41 ng/ml




#28 AR-1254-3

R.T.: 5.046 min
Delta R.T.: -0.006 min
Response: 40857213
Conc: 47.65 ng/ml m




#29 AR-1254-4

R.T.: 6.160 min
Delta R.T.: -0.006 min
Response: 53265251
Conc: 53.71 ng/ml




#29 AR-1254-4

R.T.: 5.267 min
Delta R.T.: -0.008 min
Response: 27960845
Conc: 49.94 ng/ml



#30 AR-1254-5

R.T.: 6.570 min
Delta R.T.: -0.005 min
Response: 56631961
Conc: 53.95 ng/ml



#30 AR-1254-5

R.T.: 5.664 min
Delta R.T.: -0.008 min
Response: 40466876
Conc: 49.93 ng/ml