

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ030219\
 Data File : PQ037710.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Mar 2019 11:21
 Operator : SM\SJ
 Sample : K1649-02
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 TRE-FLOOR-SLAB

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 01 21:39:30 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ022219.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Feb 23 04:01:25 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.219	3.652	116.4E6	59484214	18.572	17.803
2)	SA Decachlor...	9.676	8.541	95436338	45600297	18.487	19.772
Target Compounds							
3)	L1 AR-1016-1	5.363	4.727	13602793	2140219	73.004	20.664 #
4)	L1 AR-1016-2	5.444	4.727	4570618	2140219	16.086	13.811
5)	L1 AR-1016-3	5.444	4.901	4570618	1693259	27.516	21.851
6)	L1 AR-1016-4	5.544	4.941	4205578	4392897	30.523	69.059 #
7)	L1 AR-1016-5	5.826	5.150	6999820	3162323	51.304	39.242
8)	L2 AR-1221-1	4.488	3.908	9497057	3702152	90.288	65.041 #
9)	L2 AR-1221-2	4.520	3.908	2057416	3702152	28.354	91.938 #
10)	L2 AR-1221-3	4.603	4.015	6945541	1703950	26.708	12.008 #
11)	L3 AR-1232-1	4.603	4.015	6945541	1703950	64.880	28.272 #
12)	L3 AR-1232-2	5.100	4.727	2596683	2140219	47.078	32.625 #
13)	L3 AR-1232-3	5.444	4.901	4570618	1693259	38.713	52.617 #
14)	L3 AR-1232-4	5.544	4.980	4205578	1518314	74.811	53.795 #
15)	L3 AR-1232-5	5.667	5.150	4772386	3162323	113.354	104.135
16)	L4 AR-1242-1	5.363	4.727	13602793	2140219	92.544	26.200 #
17)	L4 AR-1242-2	5.444	4.727	4570618	2140219	20.217	17.414
18)	L4 AR-1242-3	5.444	4.901	4570618	1693259	34.500	27.332
19)	L4 AR-1242-4	5.544	4.980	4205578	1518314	38.561	25.575 #
20)	L4 AR-1242-5	6.262	5.494	12823073	17109484	106.791	237.795 #
21)	L5 AR-1248-1	5.363	4.727	13602793	2140219	134.917	37.156 #
22)	L5 AR-1248-2	5.667	4.941	4772386	4392897	34.491	59.193 #
23)	L5 AR-1248-3	5.826	4.980	6999820	1518314	45.192	20.325 #
24)	L5 AR-1248-4	6.262	5.150	12823073	3162323	71.648	34.967 #
25)	L5 AR-1248-5	6.262	5.534	12823073	5506708	76.102	61.240
26)	L6 AR-1254-1	6.223	5.494	15642513	17109484	75.975	109.075 #
27)	L6 AR-1254-2	6.458	5.637	23985838	16406692	74.749	123.272 #
28)	L6 AR-1254-3	6.820	6.040	37058907	36430324	112.698	173.266 #
29)	L6 AR-1254-4	7.090	6.262	5825236	15843076	22.247	102.088 #
30)	L6 AR-1254-5	7.496	6.669	14372897	36375062	57.504	212.073 #
31)	L7 AR-1260-1	6.926	6.158	62918235	32203971	224.869	204.557
32)	L7 AR-1260-2	7.231	6.347	17734956	35330512	49.698	185.116 #
33)	L7 AR-1260-3	7.589	6.496	25414970	31015483	109.603	178.617 #
34)	L7 AR-1260-4	7.761	6.959	44586054	19054566	180.890	166.561
35)	L7 AR-1260-5	8.109	7.202	8360849	33620753	15.849	119.995 #
36)	L8 AR-1262-1	7.589	6.701	25414970	28256182	143.771	121.831
37)	L8 AR-1262-2	8.109	7.202	8360849	33620753	10.124	81.303 #
38)	L8 AR-1262-3	8.378	7.475	3324598	7896698	9.468	48.521 #
39)	L8 AR-1262-4	8.432	7.542	19949800	20522809	48.725	70.330 #
40)	L8 AR-1262-5	9.036	8.036	8267296	4148742	30.890	34.267
41)	L9 AR-1268-1	8.378	7.475	3324598	7896698	3.575	18.386 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ030219\
 Data File : PQ037710.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Mar 2019 11:21
 Operator : SM\SJ
 Sample : K1649-02
 Misc :
 ALS Vial : 10 Sample Multiplier: 1


Instrument :
 ECD_Q
 ClientSampleId :
 TRE-FLOOR-SLAB

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 01 21:39:30 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ022219.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Feb 23 04:01:25 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.432	7.542	19949800	20522809	23.611	52.762 #
43) L9	AR-1268-3	8.629	7.735	2059034	1382004	2.744	4.285 #
44) L9	AR-1268-4	9.036	8.036	8267296	4148742	31.771	34.110
45) L9	AR-1268-5	9.394	8.309	6688578	3395577	3.258	3.639


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.



#1 Tetrachloro-m-xylene


R.T.: 4.219 min
Delta R.T.: -0.027 min
Response: 116357691
Conc: 18.57 ng/ml

Instrument :
ECD_Q
ClientSampleId :
TRE-FLOOR-SLAB




#1 Tetrachloro-m-xylene

R.T.: 3.652 min
Delta R.T.: -0.009 min
Response: 59484214
Conc: 17.80 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.676 min
Delta R.T.: -0.057 min
Response: 95436338
Conc: 18.49 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.541 min
Delta R.T.: -0.019 min
Response: 45600297
Conc: 19.77 ng/ml



#3 AR-1016-1


R.T.: 5.363 min
Delta R.T.: -0.029 min
Response: 13602793
Conc: 73.00 ng/ml

Instrument :
ECD_Q
ClientSampleId :
TRE-FLOOR-SLAB




#3 AR-1016-1

R.T.: 4.727 min
Delta R.T.: 0.005 min
Response: 2140219
Conc: 20.66 ng/ml




#4 AR-1016-2

R.T.: 5.444 min
Delta R.T.: 0.030 min
Response: 4570618
Conc: 16.09 ng/ml



#4 AR-1016-2


R.T.: 4.727 min
Delta R.T.: -0.012 min
Response: 2140219
Conc: 13.81 ng/ml



#5 AR-1016-3


R.T.: 5.444 min
Delta R.T.: -0.031 min
Response: 4570618
Conc: 27.52 ng/ml

Instrument :
ECD_Q
ClientSampleId :
TRE-FLOOR-SLAB




#5 AR-1016-3

R.T.: 4.901 min
Delta R.T.: -0.013 min
Response: 1693259
Conc: 21.85 ng/ml




#6 AR-1016-4

R.T.: 5.544 min
Delta R.T.: -0.030 min
Response: 4205578
Conc: 30.52 ng/ml



#6 AR-1016-4


R.T.: 4.941 min
Delta R.T.: -0.012 min
Response: 4392897
Conc: 69.06 ng/ml



#7 AR-1016-5


R.T.: 5.826 min
Delta R.T.: -0.032 min
Response: 6999820
Conc: 51.30 ng/ml

Instrument :
ECD_Q
ClientSampleId :
TRE-FLOOR-SLAB




#7 AR-1016-5

R.T.: 5.150 min
Delta R.T.: -0.013 min
Response: 3162323
Conc: 39.24 ng/ml




#8 AR-1221-1

R.T.: 4.488 min
Delta R.T.: 0.032 min
Response: 9497057
Conc: 90.29 ng/ml



#8 AR-1221-1


R.T.: 3.908 min
Delta R.T.: 0.041 min
Response: 3702152
Conc: 65.04 ng/ml



#9 AR-1221-2


R.T.: 4.520 min
Delta R.T.: -0.022 min
Response: 2057416
Conc: 28.35 ng/ml

Instrument :
ECD_Q
ClientSampleId :
TRE-FLOOR-SLAB




#9 AR-1221-2

R.T.: 3.908 min
Delta R.T.: -0.045 min
Response: 3702152
Conc: 91.94 ng/ml




#10 AR-1221-3

R.T.: 4.603 min
Delta R.T.: -0.015 min
Response: 6945541
Conc: 26.71 ng/ml



#10 AR-1221-3


R.T.: 4.015 min
Delta R.T.: -0.012 min
Response: 1703950
Conc: 12.01 ng/ml



#11 AR-1232-1


R.T.: 4.603 min
Delta R.T.: -0.015 min
Response: 6945541
Conc: 64.88 ng/ml

Instrument :
ECD_Q
ClientSampleId :
TRE-FLOOR-SLAB




#11 AR-1232-1

R.T.: 4.015 min
Delta R.T.: -0.012 min
Response: 1703950
Conc: 28.27 ng/ml




#12 AR-1232-2

R.T.: 5.100 min
Delta R.T.: -0.031 min
Response: 2596683
Conc: 47.08 ng/ml



#12 AR-1232-2


R.T.: 4.727 min
Delta R.T.: -0.012 min
Response: 2140219
Conc: 32.62 ng/ml



#13 AR-1232-3


R.T.: 5.444 min
Delta R.T.: 0.030 min
Response: 4570618
Conc: 38.71 ng/ml

Instrument :
ECD_Q
ClientSampleId :
TRE-FLOOR-SLAB




#13 AR-1232-3

R.T.: 4.901 min
Delta R.T.: -0.012 min
Response: 1693259
Conc: 52.62 ng/ml




#14 AR-1232-4

R.T.: 5.544 min
Delta R.T.: -0.030 min
Response: 4205578
Conc: 74.81 ng/ml



#14 AR-1232-4


R.T.: 4.980 min
Delta R.T.: -0.013 min
Response: 1518314
Conc: 53.79 ng/ml



#15 AR-1232-5


R.T.: 5.667 min
Delta R.T.: 0.012 min
Response: 4772386
Conc: 113.35 ng/ml

Instrument :
ECD_Q
ClientSampleId :
TRE-FLOOR-SLAB




#15 AR-1232-5

R.T.: 5.150 min
Delta R.T.: -0.012 min
Response: 3162323
Conc: 104.13 ng/ml




#16 AR-1242-1

R.T.: 5.363 min
Delta R.T.: -0.029 min
Response: 13602793
Conc: 92.54 ng/ml



#16 AR-1242-1


R.T.: 4.727 min
Delta R.T.: 0.005 min
Response: 2140219
Conc: 26.20 ng/ml



#17 AR-1242-2


R.T.: 5.444 min
Delta R.T.: 0.032 min
Response: 4570618
Conc: 20.22 ng/ml

Instrument :
ECD_Q
ClientSampleId :
TRE-FLOOR-SLAB




#17 AR-1242-2

R.T.: 4.727 min
Delta R.T.: -0.011 min
Response: 2140219
Conc: 17.41 ng/ml




#18 AR-1242-3

R.T.: 5.444 min
Delta R.T.: -0.029 min
Response: 4570618
Conc: 34.50 ng/ml



#18 AR-1242-3


R.T.: 4.901 min
Delta R.T.: -0.012 min
Response: 1693259
Conc: 27.33 ng/ml



#19 AR-1242-4


R.T.: 5.544 min
Delta R.T.: -0.029 min
Response: 4205578
Conc: 38.56 ng/ml

Instrument :
ECD_Q
ClientSampleId :
TRE-FLOOR-SLAB




#19 AR-1242-4

R.T.: 4.980 min
Delta R.T.: -0.013 min
Response: 1518314
Conc: 25.57 ng/ml




#20 AR-1242-5

R.T.: 6.262 min
Delta R.T.: -0.031 min
Response: 12823073
Conc: 106.79 ng/ml



#20 AR-1242-5


R.T.: 5.494 min
Delta R.T.: -0.013 min
Response: 17109484
Conc: 237.80 ng/ml



#21 AR-1248-1


R.T.: 5.363 min
Delta R.T.: -0.029 min
Response: 13602793
Conc: 134.92 ng/ml

Instrument :
ECD_Q
ClientSampleId :
TRE-FLOOR-SLAB




#21 AR-1248-1

R.T.: 4.727 min
Delta R.T.: 0.005 min
Response: 2140219
Conc: 37.16 ng/ml




#22 AR-1248-2

R.T.: 5.667 min
Delta R.T.: 0.013 min
Response: 4772386
Conc: 34.49 ng/ml



#22 AR-1248-2


R.T.: 4.941 min
Delta R.T.: -0.011 min
Response: 4392897
Conc: 59.19 ng/ml



#23 AR-1248-3


R.T.: 5.826 min
Delta R.T.: -0.031 min
Response: 6999820
Conc: 45.19 ng/ml

Instrument :
ECD_Q
ClientSampleId :
TRE-FLOOR-SLAB




#23 AR-1248-3

R.T.: 4.980 min
Delta R.T.: -0.012 min
Response: 1518314
Conc: 20.32 ng/ml




#24 AR-1248-4

R.T.: 6.262 min
Delta R.T.: 0.006 min
Response: 12823073
Conc: 71.65 ng/ml



#24 AR-1248-4


R.T.: 5.150 min
Delta R.T.: -0.012 min
Response: 3162323
Conc: 34.97 ng/ml



#25 AR-1248-5


R.T.: 6.262 min
Delta R.T.: -0.031 min
Response: 12823073
Conc: 76.10 ng/ml

Instrument :
ECD_Q
ClientSampleId :
TRE-FLOOR-SLAB




#25 AR-1248-5

R.T.: 5.534 min
Delta R.T.: -0.012 min
Response: 5506708
Conc: 61.24 ng/ml




#26 AR-1254-1

R.T.: 6.223 min
Delta R.T.: -0.002 min
Response: 15642513
Conc: 75.98 ng/ml



#26 AR-1254-1


R.T.: 5.494 min
Delta R.T.: -0.012 min
Response: 17109484
Conc: 109.08 ng/ml



#27 AR-1254-2


R.T.: 6.458 min
Delta R.T.: 0.020 min
Response: 23985838
Conc: 74.75 ng/ml

Instrument :
ECD_Q
ClientSampleId :
TRE-FLOOR-SLAB




#27 AR-1254-2

R.T.: 5.637 min
Delta R.T.: -0.013 min
Response: 16406692
Conc: 123.27 ng/ml




#28 AR-1254-3

R.T.: 6.820 min
Delta R.T.: 0.015 min
Response: 37058907
Conc: 112.70 ng/ml



#28 AR-1254-3


R.T.: 6.040 min
Delta R.T.: -0.009 min
Response: 36430324
Conc: 173.27 ng/ml



#29 AR-1254-4


R.T.: 7.090 min
Delta R.T.: 0.000 min
Response: 5825236
Conc: 22.25 ng/ml

Instrument :
ECD_Q
ClientSampleId :
TRE-FLOOR-SLAB




#29 AR-1254-4

R.T.: 6.262 min
Delta R.T.: -0.013 min
Response: 15843076
Conc: 102.09 ng/ml




#30 AR-1254-5

R.T.: 7.496 min
Delta R.T.: -0.005 min
Response: 14372897
Conc: 57.50 ng/ml



#30 AR-1254-5


R.T.: 6.669 min
Delta R.T.: -0.016 min
Response: 36375062
Conc: 212.07 ng/ml



#31 AR-1260-1


R.T.: 6.926 min
Delta R.T.: -0.035 min
Response: 62918235
Conc: 224.87 ng/ml

Instrument :
ECD_Q
ClientSampleId :
TRE-FLOOR-SLAB




#31 AR-1260-1

R.T.: 6.158 min
Delta R.T.: -0.015 min
Response: 32203971
Conc: 204.56 ng/ml




#32 AR-1260-2

R.T.: 7.231 min
Delta R.T.: 0.016 min
Response: 17734956
Conc: 49.70 ng/ml



#32 AR-1260-2


R.T.: 6.347 min
Delta R.T.: -0.014 min
Response: 35330512
Conc: 185.12 ng/ml



#33 AR-1260-3


R.T.: 7.589 min
Delta R.T.: 0.022 min
Response: 25414970
Conc: 109.60 ng/ml

Instrument :
ECD_Q
ClientSampleId :
TRE-FLOOR-SLAB




#33 AR-1260-3

R.T.: 6.496 min
Delta R.T.: -0.015 min
Response: 31015483
Conc: 178.62 ng/ml




#34 AR-1260-4

R.T.: 7.761 min
Delta R.T.: -0.035 min
Response: 44586054
Conc: 180.89 ng/ml



#34 AR-1260-4


R.T.: 6.959 min
Delta R.T.: -0.015 min
Response: 19054566
Conc: 166.56 ng/ml



#35 AR-1260-5


R.T.: 8.109 min
Delta R.T.: 0.005 min
Response: 8360849
Conc: 15.85 ng/ml

Instrument :
ECD_Q
ClientSampleId :
TRE-FLOOR-SLAB




#35 AR-1260-5

R.T.: 7.202 min
Delta R.T.: -0.015 min
Response: 33620753
Conc: 119.99 ng/ml




#36 AR-1262-1

R.T.: 7.589 min
Delta R.T.: -0.034 min
Response: 25414970
Conc: 143.77 ng/ml



#36 AR-1262-1


R.T.: 6.701 min
Delta R.T.: -0.014 min
Response: 28256182
Conc: 121.83 ng/ml



#37 AR-1262-2


R.T.: 8.109 min
Delta R.T.: 0.008 min
Response: 8360849
Conc: 10.12 ng/ml

Instrument :
ECD_Q
ClientSampleId :
TRE-FLOOR-SLAB




#37 AR-1262-2

R.T.: 7.202 min
Delta R.T.: -0.014 min
Response: 33620753
Conc: 81.30 ng/ml




#38 AR-1262-3

R.T.: 8.378 min
Delta R.T.: -0.004 min
Response: 3324598
Conc: 9.47 ng/ml



#38 AR-1262-3


R.T.: 7.475 min
Delta R.T.: -0.013 min
Response: 7896698
Conc: 48.52 ng/ml



#39 AR-1262-4


R.T.: 8.432 min
Delta R.T.: -0.037 min
Response: 19949800
Conc: 48.73 ng/ml

Instrument :
ECD_Q
ClientSampleId :
TRE-FLOOR-SLAB




#39 AR-1262-4

R.T.: 7.542 min
Delta R.T.: -0.015 min
Response: 20522809
Conc: 70.33 ng/ml




#40 AR-1262-5

R.T.: 9.036 min
Delta R.T.: -0.044 min
Response: 8267296
Conc: 30.89 ng/ml



#40 AR-1262-5


R.T.: 8.036 min
Delta R.T.: -0.014 min
Response: 4148742
Conc: 34.27 ng/ml



#41 AR-1268-1


R.T.: 8.378 min
Delta R.T.: -0.004 min
Response: 3324598
Conc: 3.58 ng/ml

Instrument :
ECD_Q
ClientSampleId :
TRE-FLOOR-SLAB




#41 AR-1268-1

R.T.: 7.475 min
Delta R.T.: -0.014 min
Response: 7896698
Conc: 18.39 ng/ml




#42 AR-1268-2

R.T.: 8.432 min
Delta R.T.: -0.039 min
Response: 19949800
Conc: 23.61 ng/ml



#42 AR-1268-2


R.T.: 7.542 min
Delta R.T.: -0.015 min
Response: 20522809
Conc: 52.76 ng/ml



#43 AR-1268-3


R.T.: 8.629 min
Delta R.T.: -0.040 min
Response: 2059034
Conc: 2.74 ng/ml

Instrument :
ECD_Q
ClientSampleId :
TRE-FLOOR-SLAB




#43 AR-1268-3

R.T.: 7.735 min
Delta R.T.: -0.015 min
Response: 1382004
Conc: 4.28 ng/ml




#44 AR-1268-4

R.T.: 9.036 min
Delta R.T.: -0.043 min
Response: 8267296
Conc: 31.77 ng/ml



#44 AR-1268-4


R.T.: 8.036 min
Delta R.T.: -0.014 min
Response: 4148742
Conc: 34.11 ng/ml



#45 AR-1268-5

R.T.: 9.394 min
Delta R.T.: -0.047 min
Response: 6688578
Conc: 3.26 ng/ml

Instrument :
ECD_Q
ClientSampleId :
TRE-FLOOR-SLAB



#45 AR-1268-5

R.T.: 8.309 min
Delta R.T.: -0.015 min
Response: 3395577
Conc: 3.64 ng/ml