

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ030219\
 Data File : PQ037705.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Mar 2019 10:01
 Operator : SM\SJ
 Sample : K1595-03RE
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 TP-CBF-B11RE

Manual Integrations
 APPROVED

Sohil
 3/4/2019 11:48:00 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 01 21:35:14 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ022219.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Feb 23 04:01:25 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

1) SA Tetrachlo...	4.220	3.653	109.7E6	57705213	17.512m	17.271
2) SA Decachlor...	9.680	8.542	150.5E6	77865438	29.155m	33.761

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ030219\
Data File : PQ037705.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Mar 2019 10:01
Operator : SM\SJ
Sample : K1595-03RE
Misc :
ALS Vial : 5 Sample Multiplier: 1



Instrument :
ECD_Q
ClientSampled :
TP-CBF-B11RE


Manual Integrations
APPROVED

Sohil
3/4/2019 11:48:00 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 01 21:35:14 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ022219.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Feb 23 04:01:25 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.220 min
Delta R.T.: -0.026 min
Response: 109716796
Conc: 17.51 ng/ml m

Instrument :
ECD_Q
ClientSampleId :
TP-CBF-B11RE


Manual Integrations
APPROVED

Sohil
3/4/2019 11:48:00 AM




#1 Tetrachloro-m-xylene

R.T.: 3.653 min
Delta R.T.: -0.008 min
Response: 57705213
Conc: 17.27 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.680 min
Delta R.T.: -0.053 min
Response: 150508464
Conc: 29.16 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.542 min
Delta R.T.: -0.018 min
Response: 77865438
Conc: 33.76 ng/ml