

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ030419\
 Data File : PQ037788.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Mar 2019 00:16
 Operator : SM\SJ
 Sample : AR1242CCC400
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1242365

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 05 00:50:04 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ022619CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Feb 26 04:57:52 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.213	3.650	138.1E6	73498683	24.279	27.904
2)	SA Decachlor...	9.660	8.533	198.1E6	88415202	39.219	38.110
Target Compounds							
16)	L4 AR-1242-1	5.356	4.707	70106104	40236517	427.381m	447.931m
17)	L4 AR-1242-2	5.376	4.724	107.5E6	59945647	426.638	439.600
18)	L4 AR-1242-3	5.437	4.899	62175247	30866269	411.043	437.625
19)	L4 AR-1242-4	5.537	4.977	51346746	30970425	407.913	451.370
20)	L4 AR-1242-5	6.253	5.491	52216767	32537891	376.268	372.162

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ030419\
Data File : PQ037788.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Mar 2019 00:16
Operator : SM\SJ
Sample : AR1242CCC400
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1242365

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 05 00:50:04 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ022619CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Feb 26 04:57:52 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

