

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ030524\
 Data File : PQ065546.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Mar 2024 16:12
 Operator : YP\AJ
 Sample : AR1221ICC800
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 06 02:25:48 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ030524CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 06 02:21:06 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml



System Monitoring Compounds							
1)	SA Tetrachlo...	3.450	2.739	172.8E6	274.0E6	37.983	38.526
2)	SA Decachlor...	8.665	7.512	114.8E6	332.2E6	74.397	73.193
Target Compounds							
8)	L2 AR-1221-1	3.652	2.941	35961169	63119247	739.867	750.306
9)	L2 AR-1221-2	3.732	3.017	24766157	46244224	756.454	755.987
10)	L2 AR-1221-3	3.803	3.084	80566764	142.6E6	735.070	731.543


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ030524\
Data File : PQ065546.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Mar 2024 16:12
Operator : YP\AJ
Sample : AR1221ICC800
Misc :
ALS Vial : 11 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 06 02:25:48 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ030524CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 06 02:21:06 2024
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.450 min
Delta R.T.: 0.000 min
Response: 172755802
Conc: 37.98 ng/ml




#1 Tetrachloro-m-xylene

R.T.: 2.739 min
Delta R.T.: 0.000 min
Response: 274010139
Conc: 38.53 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.665 min
Delta R.T.: -0.003 min
Response: 114832650
Conc: 74.40 ng/ml




#2 Decachlorobiphenyl

R.T.: 7.512 min
Delta R.T.: 0.000 min
Response: 332154563
Conc: 73.19 ng/ml




#8 AR-1221-1

R.T.: 3.652 min
Delta R.T.: 0.000 min
Response: 35961169
Conc: 739.87 ng/ml




#8 AR-1221-1

R.T.: 2.941 min
Delta R.T.: 0.000 min
Response: 63119247
Conc: 750.31 ng/ml




#9 AR-1221-2

R.T.: 3.732 min
Delta R.T.: 0.000 min
Response: 24766157
Conc: 756.45 ng/ml




#9 AR-1221-2

R.T.: 3.017 min
Delta R.T.: 0.000 min
Response: 46244224
Conc: 755.99 ng/ml



#10 AR-1221-3

R.T.: 3.803 min
Delta R.T.: -0.001 min
Response: 80566764
Conc: 735.07 ng/ml



#10 AR-1221-3

R.T.: 3.084 min
Delta R.T.: 0.000 min
Response: 142631275
Conc: 731.54 ng/ml