

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ030524\
 Data File : PQ065560.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Mar 2024 20:16
 Operator : YP\AJ
 Sample : AR1248ICC400
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 06 03:55:49 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ030524CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 06 03:54:29 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml



System Monitoring Compounds							
1)	SA Tetrachlo...	3.444	2.738	86898817	147.9E6	20.000	20.000
2)	SA Decachlor...	8.656	7.508	62417625	181.5E6	40.000	40.000
Target Compounds							
21)	L5 AR-1248-1	4.555	3.738	35083014	63138017	400.000	400.000
22)	L5 AR-1248-2	4.821	3.962	47751356	91903969	400.000	400.000
23)	L5 AR-1248-3	5.013	3.998	56958675	87855981	400.000	400.000
24)	L5 AR-1248-4	5.409	4.156	63205345	108.7E6	400.000	400.000
25)	L5 AR-1248-5	5.445	4.530	62913911	106.6E6	400.000	400.000


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ030524\
Data File : PQ065560.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Mar 2024 20:16
Operator : YP\AJ
Sample : AR1248ICC400
Misc :
ALS Vial : 25 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 06 03:55:49 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ030524CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 06 03:54:29 2024
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.444 min
Delta R.T.: 0.000 min
Response: 86898817
Conc: 20.00 ng/ml




#1 Tetrachloro-m-xylene

R.T.: 2.738 min
Delta R.T.: 0.000 min
Response: 147851094
Conc: 20.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.656 min
Delta R.T.: 0.000 min
Response: 62417625
Conc: 40.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 7.508 min
Delta R.T.: 0.000 min
Response: 181535513
Conc: 40.00 ng/ml




#21 AR-1248-1

R.T.: 4.555 min
Delta R.T.: 0.000 min
Response: 35083014
Conc: 400.00 ng/ml




#21 AR-1248-1

R.T.: 3.738 min
Delta R.T.: 0.000 min
Response: 63138017
Conc: 400.00 ng/ml




#22 AR-1248-2

R.T.: 4.821 min
Delta R.T.: 0.000 min
Response: 47751356
Conc: 400.00 ng/ml




#22 AR-1248-2

R.T.: 3.962 min
Delta R.T.: 0.000 min
Response: 91903969
Conc: 400.00 ng/ml




#23 AR-1248-3

R.T.: 5.013 min
Delta R.T.: 0.000 min
Response: 56958675
Conc: 400.00 ng/ml




#23 AR-1248-3

R.T.: 3.998 min
Delta R.T.: 0.000 min
Response: 87855981
Conc: 400.00 ng/ml




#24 AR-1248-4

R.T.: 5.409 min
Delta R.T.: 0.000 min
Response: 63205345
Conc: 400.00 ng/ml




#24 AR-1248-4

R.T.: 4.156 min
Delta R.T.: 0.000 min
Response: 108742542
Conc: 400.00 ng/ml



#25 AR-1248-5

R.T.: 5.445 min
Delta R.T.: 0.000 min
Response: 62913911
Conc: 400.00 ng/ml



#25 AR-1248-5

R.T.: 4.530 min
Delta R.T.: 0.000 min
Response: 106587855
Conc: 400.00 ng/ml