

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ030524\
 Data File : PQ065562.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Mar 2024 20:51
 Operator : YP\AJ
 Sample : AR1248ICC1600
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 06 03:56:33 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ030524CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 06 03:54:29 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml



System Monitoring Compounds						
1) SA Tetrachlo...	3.450	2.739	318.4E6	540.0E6	73.290	73.050
2) SA Decachlor...	8.662	7.511	210.2E6	615.1E6	134.729	135.527
Target Compounds						
21) L5 AR-1248-1	4.560	3.738	122.3E6	216.2E6	1394.547	1369.681
22) L5 AR-1248-2	4.828	3.962	161.8E6	308.5E6	1355.083	1342.613
23) L5 AR-1248-3	5.019	3.999	195.9E6	297.4E6	1375.932	1354.064
24) L5 AR-1248-4	5.414	4.157	218.4E6	371.4E6	1382.143	1366.243
25) L5 AR-1248-5	5.449	4.531	218.1E6	372.3E6	1386.730	1397.175


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ030524\
Data File : PQ065562.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Mar 2024 20:51
Operator : YP\AJ
Sample : AR1248ICC1600
Misc :
ALS Vial : 27 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 06 03:56:33 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ030524CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 06 03:54:29 2024
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.450 min
Delta R.T.: 0.006 min
Response: 318440582
Conc: 73.29 ng/ml




#1 Tetrachloro-m-xylene

R.T.: 2.739 min
Delta R.T.: 0.000 min
Response: 540029034
Conc: 73.05 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.662 min
Delta R.T.: 0.006 min
Response: 210237370
Conc: 134.73 ng/ml




#2 Decachlorobiphenyl

R.T.: 7.511 min
Delta R.T.: 0.002 min
Response: 615074528
Conc: 135.53 ng/ml




#21 AR-1248-1

R.T.: 4.560 min
Delta R.T.: 0.006 min
Response: 122312276
Conc: 1394.55 ng/ml




#21 AR-1248-1

R.T.: 3.738 min
Delta R.T.: 0.000 min
Response: 216197281
Conc: 1369.68 ng/ml




#22 AR-1248-2

R.T.: 4.828 min
Delta R.T.: 0.006 min
Response: 161767685
Conc: 1355.08 ng/ml




#22 AR-1248-2

R.T.: 3.962 min
Delta R.T.: 0.000 min
Response: 308478596
Conc: 1342.61 ng/ml




#23 AR-1248-3

R.T.: 5.019 min
Delta R.T.: 0.006 min
Response: 195928227
Conc: 1375.93 ng/ml




#23 AR-1248-3

R.T.: 3.999 min
Delta R.T.: 0.000 min
Response: 297406458
Conc: 1354.06 ng/ml




#24 AR-1248-4

R.T.: 5.414 min
Delta R.T.: 0.006 min
Response: 218397092
Conc: 1382.14 ng/ml




#24 AR-1248-4

R.T.: 4.157 min
Delta R.T.: 0.000 min
Response: 371421844
Conc: 1366.24 ng/ml



#25 AR-1248-5

R.T.: 5.449 min
Delta R.T.: 0.004 min
Response: 218111499
Conc: 1386.73 ng/ml



#25 AR-1248-5

R.T.: 4.531 min
Delta R.T.: 0.000 min
Response: 372304714
Conc: 1397.17 ng/ml