

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ030524\
 Data File : PQ065588.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Mar 2024 04:14
 Operator : YP\AJ
 Sample : P1645-03
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 E0684

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 03/06/2024
 Supervised By :Ankita Jodhani 03/07/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 06 05:56:14 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ030524CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 06 05:19:07 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.444	2.739	58281470	92156269	13.609	13.191
2)	SA Decachlor...	8.653	7.506	30825193	92299811	20.398	20.486
Target Compounds							
21)	L5 AR-1248-1	4.554	3.737	171.3E6	329.5E6	1937.680	2055.934m
22)	L5 AR-1248-2	4.821	3.961	262.2E6	483.9E6	2189.507	2085.408
23)	L5 AR-1248-3	5.012	3.997	305.7E6	494.3E6	2141.465	2216.313
24)	L5 AR-1248-4	5.409	4.156	410.8E6	575.7E6	2568.652	2095.377
25)	L5 AR-1248-5	5.442	4.530	530.7E6	946.6E6	3324.462	3505.350m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ030524\
Data File : PQ065588.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Mar 2024 04:14
Operator : YP\AJ
Sample : P1645-03
Misc :
ALS Vial : 53 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
E0684

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/06/2024
Supervised By :Ankita Jodhani 03/07/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 06 05:56:14 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ030524CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 06 05:19:07 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

