

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ030724\
 Data File : PQ065632.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Mar 2024 09:57
 Operator : YP\AJ
 Sample : P1601-02
 Misc : MDL 25 PPB
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 MDL-WATER-QT1-2024-02

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 07 21:36:43 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ030524CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 06 05:19:07 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.450	2.739	75032094	119.0E6	17.521	17.035
2)	SA Decachlor...	8.666	7.511	54891917	155.7E6	36.324	34.549
Target Compounds							
3)	L1 AR-1016-1	4.560	3.738	3332883	5790417	24.046	23.485
4)	L1 AR-1016-2	4.581	3.753	4569588	8134854	22.936	22.964
5)	L1 AR-1016-3	4.638	3.913	3204876	4523688	25.408	23.592
6)	L1 AR-1016-4	4.732	3.962	2912018	4013103	28.016	24.814
7)	L1 AR-1016-5	5.021	4.156	2009581	5275925	20.067	25.776 #
8)	L2 AR-1221-1	0.000	3.007f	0	3196967	N.D.	38.038 #
9)	L2 AR-1221-2	3.731	3.007	1420350	3196967	43.560	53.321
10)	L2 AR-1221-3	3.802	3.083	2154548	3882245	19.946	19.983
11)	L3 AR-1232-1	3.802	3.083	2154548	3882245	23.703	24.116
12)	L3 AR-1232-2	4.301	3.753	2237479	8134854	45.831	50.081
13)	L3 AR-1232-3	4.581	3.913	4569588	4523688	49.506	53.210
14)	L3 AR-1232-4	4.732	3.998	2912018	4696102	60.389	63.793
15)	L3 AR-1232-5	4.828	4.156	2410541	5275925	74.796	60.930
16)	L4 AR-1242-1	4.560	3.738	3332883	5790417	28.110	27.686
17)	L4 AR-1242-2	4.581	3.753	4569588	8134854	27.378	27.570
18)	L4 AR-1242-3	4.638	3.913	3204876	4523688	30.280	28.175
19)	L4 AR-1242-4	4.732	3.998	2912018	4696102	33.287	30.918
20)	L4 AR-1242-5	5.414f	4.492	1111845	4259253	12.084	21.697 #
21)	L5 AR-1248-1	4.560	3.738	3332883	5790417	37.692	36.126
22)	L5 AR-1248-2	4.828	3.962	2410541	4013103	20.129	17.293
23)	L5 AR-1248-3	5.021	3.998	2009581	4696102	14.078	21.056 #
24)	L5 AR-1248-4	5.414	4.156	1111845	5275925	6.952	19.201 #
25)	L5 AR-1248-5	5.414f	4.492f	1111845	4259253	6.965	15.772 #
26)	L6 AR-1254-1	5.388	4.492	1833105	4259253	11.925	10.370
27)	L6 AR-1254-2	5.607	4.640	1836581	3259694	7.859	8.917
28)	L6 AR-1254-3	5.958	5.040	1362708	7024344	5.593	12.017 #
29)	L6 AR-1254-4	0.000	5.282	0	2608438	N.D.	6.868 #
30)	L6 AR-1254-5	6.659	5.653	6189663	13790035	33.885	25.523
31)	L7 AR-1260-1	6.125	5.150	4651859	10210638	25.440	24.884
32)	L7 AR-1260-2	6.387	5.340	5136397	11788564	24.041	23.513
33)	L7 AR-1260-3	6.742	5.482	3524491	11871567	22.879	25.342
34)	L7 AR-1260-4	6.960	5.943	3772868	8376084	22.309	21.533
35)	L7 AR-1260-5	7.272	6.191	7428506	21094961	22.489	22.797
36)	L8 AR-1262-1	6.742	5.696	3524491	9744944	15.875	17.318
37)	L8 AR-1262-2	7.272	6.191	7428506	21094961	20.404	20.997
38)	L8 AR-1262-3	7.552	6.469	5118471	5386151	21.269	13.052 #
39)	L8 AR-1262-4	7.609	6.527	2798996	15407484	15.250	20.523 #
40)	L8 AR-1262-5	8.129	7.026	2030331	5671469	16.631	15.630
41)	L9 AR-1268-1	7.552	6.469	5118471	5386151	12.948	4.905 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ030724\
Data File : PQ065632.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Mar 2024 09:57
Operator : YP\AJ
Sample : P1601-02
Misc : MDL 25 PPB
ALS Vial : 52 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
MDL-WATER-QT1-2024-02

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 07 21:36:43 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ030524CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 06 05:19:07 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	7.609	6.527	2798996	15407484	7.794	15.432 #
44) L9	AR-1268-4	8.129	7.026	2030331	5671469	15.770	14.878
45) L9	AR-1268-5	8.424	7.294	1086131	2367464	1.224	0.973



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ030724\
Data File : PQ065632.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Mar 2024 09:57
Operator : YP\AJ
Sample : P1601-02
Misc : MDL 25 PPB
ALS Vial : 52 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
MDL-WATER-QT1-2024-02

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 07 21:36:43 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ030524CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 06 05:19:07 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.450 min
Delta R.T.: -0.002 min
Response: 75032094
Conc: 17.52 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-WATER-QT1-2024-02




#1 Tetrachloro-m-xylene

R.T.: 2.739 min
Delta R.T.: 0.000 min
Response: 119013345
Conc: 17.03 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.666 min
Delta R.T.: -0.004 min
Response: 54891917
Conc: 36.32 ng/ml



#2 Decachlorobiphenyl


R.T.: 7.511 min
Delta R.T.: -0.003 min
Response: 155660355
Conc: 34.55 ng/ml



#3 AR-1016-1


R.T.: 4.560 min
Delta R.T.: -0.003 min
Response: 3332883
Conc: 24.05 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-WATER-QT1-2024-02




#3 AR-1016-1

R.T.: 3.738 min
Delta R.T.: 0.000 min
Response: 5790417
Conc: 23.49 ng/ml




#4 AR-1016-2

R.T.: 4.581 min
Delta R.T.: -0.003 min
Response: 4569588
Conc: 22.94 ng/ml



#4 AR-1016-2


R.T.: 3.753 min
Delta R.T.: 0.000 min
Response: 8134854
Conc: 22.96 ng/ml



#5 AR-1016-3


R.T.: 4.638 min
Delta R.T.: -0.003 min
Response: 3204876
Conc: 25.41 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-WATER-QT1-2024-02




#5 AR-1016-3

R.T.: 3.913 min
Delta R.T.: 0.000 min
Response: 4523688
Conc: 23.59 ng/ml




#6 AR-1016-4

R.T.: 4.732 min
Delta R.T.: -0.001 min
Response: 2912018
Conc: 28.02 ng/ml



#6 AR-1016-4


R.T.: 3.962 min
Delta R.T.: 0.000 min
Response: 4013103
Conc: 24.81 ng/ml



#7 AR-1016-5


R.T.: 5.021 min
Delta R.T.: -0.002 min
Response: 2009581
Conc: 20.07 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-WATER-QT1-2024-02




#7 AR-1016-5

R.T.: 4.156 min
Delta R.T.: -0.002 min
Response: 5275925
Conc: 25.78 ng/ml




#8 AR-1221-1

R.T.: 0.000 min
Exp R.T. : 3.653 min
Response: 0
Conc: N.D.



#8 AR-1221-1


R.T.: 3.007 min
Delta R.T.: 0.067 min
Response: 3196967
Conc: 38.04 ng/ml



#9 AR-1221-2


R.T.: 3.731 min
Delta R.T.: -0.001 min
Response: 1420350
Conc: 43.56 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-WATER-QT1-2024-02




#9 AR-1221-2

R.T.: 3.007 min
Delta R.T.: -0.010 min
Response: 3196967
Conc: 53.32 ng/ml




#10 AR-1221-3

R.T.: 3.802 min
Delta R.T.: 0.000 min
Response: 2154548
Conc: 19.95 ng/ml



#10 AR-1221-3


R.T.: 3.083 min
Delta R.T.: 0.000 min
Response: 3882245
Conc: 19.98 ng/ml



#11 AR-1232-1


R.T.: 3.802 min
Delta R.T.: 0.000 min
Response: 2154548
Conc: 23.70 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-WATER-QT1-2024-02




#11 AR-1232-1

R.T.: 3.083 min
Delta R.T.: -0.001 min
Response: 3882245
Conc: 24.12 ng/ml




#12 AR-1232-2

R.T.: 4.301 min
Delta R.T.: 0.000 min
Response: 2237479
Conc: 45.83 ng/ml



#12 AR-1232-2


R.T.: 3.753 min
Delta R.T.: 0.000 min
Response: 8134854
Conc: 50.08 ng/ml



#13 AR-1232-3


R.T.: 4.581 min
Delta R.T.: 0.000 min
Response: 4569588
Conc: 49.51 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-WATER-QT1-2024-02




#13 AR-1232-3

R.T.: 3.913 min
Delta R.T.: 0.000 min
Response: 4523688
Conc: 53.21 ng/ml




#14 AR-1232-4

R.T.: 4.732 min
Delta R.T.: 0.000 min
Response: 2912018
Conc: 60.39 ng/ml



#14 AR-1232-4


R.T.: 3.998 min
Delta R.T.: 0.000 min
Response: 4696102
Conc: 63.79 ng/ml



#15 AR-1232-5


R.T.: 4.828 min
Delta R.T.: 0.000 min
Response: 2410541
Conc: 74.80 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-WATER-QT1-2024-02




#15 AR-1232-5

R.T.: 4.156 min
Delta R.T.: 0.000 min
Response: 5275925
Conc: 60.93 ng/ml




#16 AR-1242-1

R.T.: 4.560 min
Delta R.T.: -0.003 min
Response: 3332883
Conc: 28.11 ng/ml



#16 AR-1242-1


R.T.: 3.738 min
Delta R.T.: 0.000 min
Response: 5790417
Conc: 27.69 ng/ml



#17 AR-1242-2


R.T.: 4.581 min
Delta R.T.: -0.003 min
Response: 4569588
Conc: 27.38 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-WATER-QT1-2024-02




#17 AR-1242-2

R.T.: 3.753 min
Delta R.T.: 0.000 min
Response: 8134854
Conc: 27.57 ng/ml




#18 AR-1242-3

R.T.: 4.638 min
Delta R.T.: -0.002 min
Response: 3204876
Conc: 30.28 ng/ml



#18 AR-1242-3


R.T.: 3.913 min
Delta R.T.: 0.000 min
Response: 4523688
Conc: 28.18 ng/ml



#19 AR-1242-4


R.T.: 4.732 min
Delta R.T.: 0.000 min
Response: 2912018
Conc: 33.29 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-WATER-QT1-2024-02




#19 AR-1242-4

R.T.: 3.998 min
Delta R.T.: -0.001 min
Response: 4696102
Conc: 30.92 ng/ml




#20 AR-1242-5

R.T.: 5.414 min
Delta R.T.: -0.038 min
Response: 1111845
Conc: 12.08 ng/ml



#20 AR-1242-5


R.T.: 4.492 min
Delta R.T.: -0.002 min
Response: 4259253
Conc: 21.70 ng/ml



#21 AR-1248-1


R.T.: 4.560 min
Delta R.T.: -0.002 min
Response: 3332883
Conc: 37.69 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-WATER-QT1-2024-02




#21 AR-1248-1

R.T.: 3.738 min
Delta R.T.: 0.000 min
Response: 5790417
Conc: 36.13 ng/ml




#22 AR-1248-2

R.T.: 4.828 min
Delta R.T.: 0.000 min
Response: 2410541
Conc: 20.13 ng/ml



#22 AR-1248-2


R.T.: 3.962 min
Delta R.T.: 0.000 min
Response: 4013103
Conc: 17.29 ng/ml



#23 AR-1248-3


R.T.: 5.021 min
Delta R.T.: 0.000 min
Response: 2009581
Conc: 14.08 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-WATER-QT1-2024-02




#23 AR-1248-3

R.T.: 3.998 min
Delta R.T.: 0.000 min
Response: 4696102
Conc: 21.06 ng/ml




#24 AR-1248-4

R.T.: 5.414 min
Delta R.T.: -0.002 min
Response: 1111845
Conc: 6.95 ng/ml



#24 AR-1248-4


R.T.: 4.156 min
Delta R.T.: 0.000 min
Response: 5275925
Conc: 19.20 ng/ml



#25 AR-1248-5


R.T.: 5.414 min
Delta R.T.: -0.036 min
Response: 1111845
Conc: 6.97 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-WATER-QT1-2024-02




#25 AR-1248-5

R.T.: 4.492 min
Delta R.T.: -0.039 min
Response: 4259253
Conc: 15.77 ng/ml




#26 AR-1254-1

R.T.: 5.388 min
Delta R.T.: 0.000 min
Response: 1833105
Conc: 11.93 ng/ml



#26 AR-1254-1


R.T.: 4.492 min
Delta R.T.: -0.001 min
Response: 4259253
Conc: 10.37 ng/ml



#27 AR-1254-2


R.T.: 5.607 min
Delta R.T.: 0.002 min
Response: 1836581
Conc: 7.86 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-WATER-QT1-2024-02




#27 AR-1254-2

R.T.: 4.640 min
Delta R.T.: 0.000 min
Response: 3259694
Conc: 8.92 ng/ml




#28 AR-1254-3

R.T.: 5.958 min
Delta R.T.: -0.002 min
Response: 1362708
Conc: 5.59 ng/ml



#28 AR-1254-3


R.T.: 5.040 min
Delta R.T.: 0.018 min
Response: 7024344
Conc: 12.02 ng/ml



#29 AR-1254-4


R.T.: 0.000 min
Exp R.T.: 6.246 min
Response: 0
Conc: N.D.

Instrument :
ECD_Q
ClientSampleId :
MDL-WATER-QT1-2024-02




#29 AR-1254-4

R.T.: 5.282 min
Delta R.T.: 0.033 min
Response: 2608438
Conc: 6.87 ng/ml




#30 AR-1254-5

R.T.: 6.659 min
Delta R.T.: 0.000 min
Response: 6189663
Conc: 33.88 ng/ml



#30 AR-1254-5


R.T.: 5.653 min
Delta R.T.: -0.002 min
Response: 13790035
Conc: 25.52 ng/ml



#31 AR-1260-1


R.T.: 6.125 min
Delta R.T.: -0.005 min
Response: 4651859
Conc: 25.44 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-WATER-QT1-2024-02




#31 AR-1260-1

R.T.: 5.150 min
Delta R.T.: 0.000 min
Response: 10210638
Conc: 24.88 ng/ml




#32 AR-1260-2

R.T.: 6.387 min
Delta R.T.: -0.002 min
Response: 5136397
Conc: 24.04 ng/ml



#32 AR-1260-2


R.T.: 5.340 min
Delta R.T.: 0.000 min
Response: 11788564
Conc: 23.51 ng/ml



#33 AR-1260-3


R.T.: 6.742 min
Delta R.T.: -0.004 min
Response: 3524491
Conc: 22.88 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-WATER-QT1-2024-02




#33 AR-1260-3

R.T.: 5.482 min
Delta R.T.: -0.001 min
Response: 11871567
Conc: 25.34 ng/ml




#34 AR-1260-4

R.T.: 6.960 min
Delta R.T.: -0.003 min
Response: 3772868
Conc: 22.31 ng/ml



#34 AR-1260-4


R.T.: 5.943 min
Delta R.T.: -0.003 min
Response: 8376084
Conc: 21.53 ng/ml



#35 AR-1260-5


R.T.: 7.272 min
Delta R.T.: -0.003 min
Response: 7428506
Conc: 22.49 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-WATER-QT1-2024-02




#35 AR-1260-5

R.T.: 6.191 min
Delta R.T.: -0.001 min
Response: 21094961
Conc: 22.80 ng/ml




#36 AR-1262-1

R.T.: 6.742 min
Delta R.T.: 0.000 min
Response: 3524491
Conc: 15.88 ng/ml



#36 AR-1262-1


R.T.: 5.696 min
Delta R.T.: 0.000 min
Response: 9744944
Conc: 17.32 ng/ml



#37 AR-1262-2


R.T.: 7.272 min
Delta R.T.: 0.003 min
Response: 7428506
Conc: 20.40 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-WATER-QT1-2024-02




#37 AR-1262-2

R.T.: 6.191 min
Delta R.T.: 0.000 min
Response: 21094961
Conc: 21.00 ng/ml




#38 AR-1262-3

R.T.: 7.552 min
Delta R.T.: 0.004 min
Response: 5118471
Conc: 21.27 ng/ml



#38 AR-1262-3


R.T.: 6.469 min
Delta R.T.: 0.000 min
Response: 5386151
Conc: 13.05 ng/ml



#39 AR-1262-4


R.T.: 7.609 min
Delta R.T.: -0.013 min
Response: 2798996
Conc: 15.25 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-WATER-QT1-2024-02




#39 AR-1262-4

R.T.: 6.527 min
Delta R.T.: -0.004 min
Response: 15407484
Conc: 20.52 ng/ml




#40 AR-1262-5

R.T.: 8.129 min
Delta R.T.: 0.002 min
Response: 2030331
Conc: 16.63 ng/ml



#40 AR-1262-5


R.T.: 7.026 min
Delta R.T.: 0.001 min
Response: 5671469
Conc: 15.63 ng/ml



#41 AR-1268-1


R.T.: 7.552 min
Delta R.T.: 0.009 min
Response: 5118471
Conc: 12.95 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-WATER-QT1-2024-02




#41 AR-1268-1

R.T.: 6.469 min
Delta R.T.: 0.000 min
Response: 5386151
Conc: 4.91 ng/ml




#42 AR-1268-2

R.T.: 7.609 min
Delta R.T.: -0.014 min
Response: 2798996
Conc: 7.79 ng/ml



#42 AR-1268-2


R.T.: 6.527 min
Delta R.T.: -0.005 min
Response: 15407484
Conc: 15.43 ng/ml



#44 AR-1268-4


R.T.: 8.129 min
Delta R.T.: 0.000 min
Response: 2030331
Conc: 15.77 ng/ml

Instrument :
ECD_Q
ClientSampleId :
MDL-WATER-QT1-2024-02




#44 AR-1268-4

R.T.: 7.026 min
Delta R.T.: 0.002 min
Response: 5671469
Conc: 14.88 ng/ml



#45 AR-1268-5

R.T.: 8.424 min
Delta R.T.: -0.003 min
Response: 1086131
Conc: 1.22 ng/ml



#45 AR-1268-5

R.T.: 7.294 min
Delta R.T.: 0.000 min
Response: 2367464
Conc: 0.97 ng/ml