

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ030922\
 Data File : PQ056391.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Mar 2022 15:08
 Operator : YP\AJ
 Sample : N1645-07
 Misc : AR1232 LOD 25 PPB
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 LOD-MDL-WATER-01-QT1-2022

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 03/11/2022
 Supervised By :Ankita Jodhani 03/10/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 10 02:34:05 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ030822.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 09 03:41:18 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.677	3.846	271.1E6	253.8E6	20.670	20.220
2) SA Decachlor...	10.586	8.838	341.0E6	187.9E6	21.468	19.744
Target Compounds						
11) L3 AR-1232-1	5.061	4.222	7325968	7215921	30.206	35.366
12) L3 AR-1232-2	5.589	4.946	4192794	6700996	29.383m	25.570
13) L3 AR-1232-3	5.886	5.120	9281314	2864967	34.554m	25.510 #
14) L3 AR-1232-4	6.052	5.197	4298666	3153856	29.931m	28.809
15) L3 AR-1232-5	6.131	5.367	2398633	3172753	22.827m	27.220

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ030922\
Data File : PQ056391.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Mar 2022 15:08
Operator : YP\AJ
Sample : N1645-07
Misc : AR1232 LOD 25 PPB
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT1-2022

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/11/2022
Supervised By :Ankita Jodhani 03/10/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 10 02:34:05 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ030822.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 09 03:41:18 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

