

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ030924\
 Data File : PQ065693.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Mar 2024 23:50
 Operator : YP\AJ
 Sample : P1707-04 10X
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 HIA754R-1-2

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 03/11/2024
 Supervised By :Ankita Jodhani 03/11/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 09 00:45:17 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ022624.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Feb 27 08:43:13 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.449	2.738	5623386	10542674	1.418m	1.551
2) SA Decachlor...	8.662	7.508	4705910	7640154	1.927	0.988 #
Target Compounds						
26) L6 AR-1254-1	5.387	4.492	5806896	14393727	42.404	38.168
27) L6 AR-1254-2	5.604	4.639	10720258	12336195	51.561	36.200 #
28) L6 AR-1254-3	5.959	5.020	6964486	17419901	33.020	32.911
29) L6 AR-1254-4	6.244	5.247	5654804	10144249	36.229	30.558m
30) L6 AR-1254-5	6.659	5.651	6048245	13313232	35.561m	25.320 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ030924\
Data File : PQ065693.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Mar 2024 23:50
Operator : YP\AJ
Sample : P1707-04 10X
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
HIA754R-1-2

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/11/2024
Supervised By :Ankita Jodhani 03/11/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 09 00:45:17 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ022624.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Feb 27 08:43:13 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

