

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ031021\
 Data File : PQ052493.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Mar 2021 00:41
 Operator : DD\AJ
 Sample : M1458-07
 Misc : AR1221 LOD 25 PPB
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 LOD-MDL-WATER-01-QT1-2021

Manual Integrations
 APPROVED

Ankita
 3/10/2021 2:40:03 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 10 02:52:58 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ031021.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 10 02:40:50 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	5.226	4.242	450.1E6	155.6E6	17.722	17.109
2) SA Decachlor...	11.413	9.585	457.7E6	173.6E6	18.361	17.894
Target Compounds						
8) L2 AR-1221-1	5.485	4.497	16896174	2054867	67.701m	22.764m#
9) L2 AR-1221-2	5.578	4.597	8482203	4605224	46.827m	69.336 #
10) L2 AR-1221-3	5.662	4.689	16236373	5335168	28.184m	24.080

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ031021\
Data File : PQ052493.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Mar 2021 00:41
Operator : DD\AJ
Sample : M1458-07
Misc : AR1221 LOD 25 PPB
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT1-2021

Manual Integrations
APPROVED

Ankita
3/10/2021 2:40:03 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 10 02:52:58 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ031021.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 10 02:40:50 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

