

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ031022\
 Data File : PQ056479.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Mar 2022 14:22
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 10 22:59:55 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ030822.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 09 03:41:18 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.677	3.845	650.1E6	612.9E6	49.563	48.819
2)	SA Decachlor...	10.583	8.836	807.3E6	461.2E6	50.825	48.472
Target Compounds							
3)	L1 AR-1016-1	5.865	4.924	146.5E6	156.0E6	440.655	448.189
4)	L1 AR-1016-2	5.888	4.943	334.4E6	313.7E6	498.057	465.679
5)	L1 AR-1016-3	5.949	5.117	233.4E6	136.8E6	545.505	466.223
6)	L1 AR-1016-4	6.053	5.154	182.6E6	131.5E6	515.129	510.942
7)	L1 AR-1016-5	6.339	5.367	143.5E6	148.6E6	500.646	485.055
31)	L7 AR-1260-1	7.464	6.389	263.9E6	305.4E6	553.851	490.225
32)	L7 AR-1260-2	7.722	6.576	320.0E6	359.8E6	520.756	494.331
33)	L7 AR-1260-3	8.081	6.728	268.6E6	339.5E6	542.256	505.266
34)	L7 AR-1260-4	8.321	7.196	301.0E6	250.7E6	521.688	490.551
35)	L7 AR-1260-5	8.662	7.436	686.0E6	571.8E6	513.966	489.486

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ031022\
Data File : PQ056479.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Mar 2022 14:22
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 10 22:59:55 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ030822.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 09 03:41:18 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

