Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ031319\
Data File : PQ037960.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Mar 2019 01:42
Operator : SM\SJ
Sample : AR1254CCC400
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1254378

Manual Integrations APPROVED
Sohil
3/14/2019 9:02:00 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 14 02:23:24 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ022619CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Feb 26 04:57:52 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.201	3.645	110.5E6	59369278	19.422	22.540
2)	SA Decachlor...	9.629	8.519	189.5E6	90773556	37.522m	39.126
	Target Compounds						
26)	L6 AR-1254-1	6.168	5.482	74578688	63264256	352.135m	381.541
27)	L6 AR-1254-2	6.379	5.625	116.2E6	53866495	346.514m	378.241
28)	L6 AR-1254-3	6.748	6.022	120.5E6	83351415	347.214m	365.318m
29)	L6 AR-1254-4	7.028	6.248	91142158	62258503	329.328m	362.520
30)	L6 AR-1254-5	7.439	6.656	88698147	71846692	320.925m	371.056

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ031319\
Data File : PQ037960.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Mar 2019 01:42
Operator : SM\SJ
Sample : AR1254CCC400
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1254378

Manual Integrations APPROVED
Sohil
3/14/2019 9:02:00 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 14 02:23:24 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ022619CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Feb 26 04:57:52 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm