

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ031522\
 Data File : PQ056534.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Mar 2022 15:03
 Operator : YP\AJ
 Sample : AR1660ICC400
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR16603023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 16 06:20:22 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ031522CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 16 06:08:32 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.531	3.750	302.3E6	244.1E6	20.000	20.000
2)	SA Decachlor...	10.303	8.692	570.3E6	392.7E6	40.000	40.000
Target Compounds							
3)	L1 AR-1016-1	5.709	4.819	131.1E6	151.3E6	400.000	400.000
4)	L1 AR-1016-2	5.731	4.837	263.5E6	261.6E6	400.000	400.000
5)	L1 AR-1016-3	5.793	5.011	175.8E6	118.7E6	400.000	400.000
6)	L1 AR-1016-4	5.893	5.049	140.9E6	97918600	400.000	400.000
7)	L1 AR-1016-5	6.181	5.260	114.4E6	124.1E6	400.000	400.000
31)	L7 AR-1260-1	7.300	6.277	185.6E6	225.9E6	400.000	400.000
32)	L7 AR-1260-2	7.557	6.462	213.0E6	274.1E6	400.000	400.000
33)	L7 AR-1260-3	7.844	6.614	280.9E6	256.3E6	400.000	400.000
34)	L7 AR-1260-4	8.145	7.080	206.5E6	214.9E6	400.000	400.000
35)	L7 AR-1260-5	8.473	7.319	459.5E6	534.0E6	400.000	400.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ031522\
Data File : PQ056534.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Mar 2022 15:03
Operator : YP\AJ
Sample : AR1660ICC400
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR16603023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 16 06:20:22 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ031522CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 16 06:08:32 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

