

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ031822\
 Data File : PQ056710.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Mar 2022 11:57
 Operator : YP\AJ
 Sample : AR1248CCC400
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR12483046

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 03/21/2022
 Supervised By :mohammad ahmed 03/21/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 21 05:04:31 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ031522CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 16 09:32:52 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.531	3.748	307.6E6	247.7E6	20.509	20.402
2) SA Decachlor...	10.303	8.687	638.1E6	412.0E6	44.855	41.974
Target Compounds						
21) L5 AR-1248-1	5.709	4.816	78364104	90301838	451.467	422.356
22) L5 AR-1248-2	5.973	5.045	124.7E6	135.7E6	423.516	427.098
23) L5 AR-1248-3	6.180	5.087	152.4E6	141.9E6	419.755	424.267
24) L5 AR-1248-4	6.588	5.256	169.6E6	169.1E6	413.231	425.281
25) L5 AR-1248-5	6.627	5.645	193.6E6	172.6E6	410.256	422.499m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ031822\
Data File : PQ056710.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2022 11:57
Operator : YP\AJ
Sample : AR1248CCC400
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR12483046

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/21/2022
Supervised By :mohammad ahmed 03/21/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 21 05:04:31 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ031522CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 16 09:32:52 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

