

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_Q\Data\PQ032024\  
Data File : PQ065826.D  
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
Acq On : 20 Mar 2024 14:15  
Operator : YP\AJ  
Sample : PB159678BL  
Misc :  
ALS Vial : 7 Sample Multiplier: 1

Instrument :  
ECD\_Q  
ClientSampleId :  
PB159678BL

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Mar 20 22:52:15 2024  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_Q\Method\PQ031824.M  
Quant Title : GC EXTRACTABLES  
QLast Update : Tue Mar 19 12:55:45 2024  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 2 µl  
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

| Compound                    | RT#1  | RT#2  | Resp#1   | Resp#2  | ng/ml  | ng/ml  |
|-----------------------------|-------|-------|----------|---------|--------|--------|
| -----                       |       |       |          |         |        |        |
| System Monitoring Compounds |       |       |          |         |        |        |
| 1) SA Tetrachlo...          | 3.445 | 2.737 | 96266531 | 142.9E6 | 19.368 | 19.275 |
| 2) SA Decachlor...          | 8.662 | 7.508 | 53447860 | 148.6E6 | 18.418 | 17.777 |

Target Compounds

-----

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_Q\Data\PQ032024\  
Data File : PQ065826.D  
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
Acq On : 20 Mar 2024 14:15  
Operator : YP\AJ  
Sample : PB159678BL  
Misc :  
ALS Vial : 7 Sample Multiplier: 1

Instrument :  
ECD\_Q  
ClientSampleId :  
PB159678BL

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Mar 20 22:52:15 2024  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_Q\Method\PQ031824.M  
Quant Title : GC EXTRACTABLES  
QLast Update : Tue Mar 19 12:55:45 2024  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 2 µl  
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

