

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ032119\
Data File : PQ038044.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Mar 2019 17:58
Operator : SM\SJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 21 08:41:36 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ032019.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 20 09:51:59 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.182	3.636	119.3E6	65038021	20.510	21.058
2) SA Decachlor...	9.597	8.508	108.1E6	44555455	22.413	19.902

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ032119\
Data File : PQ038044.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Mar 2019 17:58
Operator : SM\SJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 21 08:41:36 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ032019.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 20 09:51:59 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.182 min
Delta R.T.: -0.008 min
Response: 119339963
Conc: 20.51 ng/ml

Instrument :
ECD_Q
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.636 min
Delta R.T.: -0.003 min
Response: 65038021
Conc: 21.06 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.597 min
Delta R.T.: -0.015 min
Response: 108061285
Conc: 22.41 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.508 min
Delta R.T.: -0.006 min
Response: 44555455
Conc: 19.90 ng/ml